

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW032619\
 Data File : VW009457.D
 Acq On : 26 Mar 2019 10:28
 Operator : SY/VA
 Sample : VW0326SBL01
 Misc : 5.00G/10ML/MSVOA W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VBLK97

Quant Time: Mar 27 01:44:12 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM031419S.M
 Quant Title : VOC Analysis
 QLast Update : Wed Mar 27 01:27:40 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.84	114	368488	25.00	ug/L	0.00
28) Chlorobenzene-d5	11.63	117	342755	25.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.56	152	158892	25.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.35	65	110719	25.14	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	100.56%
7) Chloroethane-d5	2.88	69	119514	23.20	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	92.80%
10) 1,1-Dichloroethene-d2	4.01	63	178232	18.73	ug/L	-0.01
Spiked Amount	25.000	Range	45 - 110	Recovery	=	74.92%
20) 2-Butanone-d5	7.08	46	108957	59.83	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	119.66%
24) Chloroform-d	7.64	84	239397	23.81	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	95.24%
26) 1,2-Dichloroethane-d4	8.30	65	147576	24.36	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	97.44%
29) Benzene-d6	8.27	84	450704	23.84	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	95.36%
33) 1,2-Dichloropropane-d6	9.27	67	136888	24.11	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	96.44%
37) Toluene-d8	10.32	98	407957	23.34	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	93.36%
38) trans-1,3-Dichloropropene-	10.57	79	62079	21.77	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	87.08%
39) 2-Hexanone-d5	10.93	63	74321	54.05	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	108.10%
48) 1,1,2,2-Tetrachloroethane-	12.69	84	135148	24.17	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	96.68%
61) 1,2-Dichlorobenzene-d4	13.86	152	149421	24.52	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	98.08%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW032619\
Data File : VW009457.D
Acq On : 26 Mar 2019 10:28
Operator : SY/VA
Sample : VW0326SBL01
Misc : 5.00G/10ML/MSVOA W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VBLK97

Quant Time: Mar 27 01:44:12 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM031419S.M
Quant Title : VOC Analysis
QLast Update : Wed Mar 27 01:27:40 2019
Response via : Initial Calibration

