

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW032619\
 Data File : VW009469.D
 Acq On : 26 Mar 2019 16:18
 Operator : SY/VA
 Sample : K2044-06RE
 Misc : 5.75G/10ML/MSVOA W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 BF5T6RE

Quant Time: Apr 08 10:58:05 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM031419S.M
 Quant Title : VOC Analysis
 QLast Update : Wed Mar 27 01:27:49 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.84	114	353715	25.00	ug/L	0.00
28) Chlorobenzene-d5	11.63	117	247699	25.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.56	152	56285	25.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.36	65	91854	21.73	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	86.92%
7) Chloroethane-d5	2.89	69	100566	20.33	ug/L	0.01
Spiked Amount	25.000	Range	30 - 150	Recovery	=	81.32%
10) 1,1-Dichloroethene-d2	4.01	63	147849	16.18	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	64.72%
20) 2-Butanone-d5	7.08	46	61188	35.00	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	70.00%
24) Chloroform-d	7.64	84	229724	23.80	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	95.20%
26) 1,2-Dichloroethane-d4	8.31	65	133953	23.04	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	92.16%
29) Benzene-d6	8.27	84	395766	28.97	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	115.88%
33) 1,2-Dichloropropane-d6	9.27	67	127256	31.02	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	124.08%#
37) Toluene-d8	10.32	98	292964	23.19	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	92.76%
38) trans-1,3-Dichloropropene-	10.57	79	44056	21.38	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	85.52%
39) 2-Hexanone-d5	10.93	63	39695	39.95	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	79.90%
48) 1,1,2,2-Tetrachloroethane-	12.69	84	87729	21.71	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	86.84%
61) 1,2-Dichlorobenzene-d4	13.86	152	47797	22.15	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	88.60%

Target Compounds

					Ovalue
13) Acetone	4.12	43	4599	3.208	ug/L 82
16) Methylene chloride	4.92	84	33882	6.569	ug/L 96
53) Ethylbenzene	11.73	91	24921	1.495	ug/L 95
56) Styrene	12.18	104	40653	3.857	ug/L 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW032619\
Data File : VW009469.D
Acq On : 26 Mar 2019 16:18
Operator : SY/VA
Sample : K2044-06RE
Misc : 5.75G/10ML/MSVOA W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BF5T6RE

Quant Time: Apr 08 10:58:05 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM031419S.M
Quant Title : VOC Analysis
QLast Update : Wed Mar 27 01:27:49 2019
Response via : Initial Calibration

