

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW032619\
 Data File : VW009490.D
 Acq On : 27 Mar 2019 03:43
 Operator : SY/VA
 Sample : VIBLK
 Misc : 5.00G/10ML/MSVOA W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VIBLK17

Manual Integrations
 APPROVED

MMDadoda
 3/30/2019 12:37:49 PM

Quant Time: Mar 27 09:28:38 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM031419S.M
 Quant Title : VOC Analysis
 QLast Update : Wed Mar 27 08:01:36 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.84	114	354676	25.00	ug/L	0.00
28) Chlorobenzene-d5	11.63	117	350786	25.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.56	152	161048	25.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.36	65	90205	21.28	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	85.12%
7) Chloroethane-d5	2.89	69	105910	21.36	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	85.44%
10) 1,1-Dichloroethene-d2	4.01	63	155394	16.96	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	67.84%
20) 2-Butanone-d5	7.08	46	95821	54.67	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	109.34%
24) Chloroform-d	7.64	84	240486	24.85	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	99.40%
26) 1,2-Dichloroethane-d4	8.31	65	155369	26.65	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	106.60%
29) Benzene-d6	8.27	84	439681	22.73	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	90.92%
33) 1,2-Dichloropropane-d6	9.27	67	140280	24.14	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	96.56%
37) Toluene-d8	10.32	98	396018	22.14	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	88.56%
38) trans-1,3-Dichloropropene-	10.58	79	60322	20.67	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	82.68%
39) 2-Hexanone-d5	10.93	63	69689	49.52	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	99.04%
48) 1,1,2,2-Tetrachloroethane-	12.69	84	143282	25.04	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	100.16%
61) 1,2-Dichlorobenzene-d4	13.85	152	154104	24.95	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	99.80%

Target Compounds				Ovalue
13) Acetone	4.13	43	2887m	2.008 ug/L
16) Methylene chloride	4.92	84	10381m	2.007 ug/L

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW032619\
Data File : VW009490.D
Acq On : 27 Mar 2019 03:43
Operator : SY/VA
Sample : VIBLK
Misc : 5.00G/10ML/MSVOA W/SOIL
ALS Vial : 1 Sample Multiplier: 1


Instrument :
MSVOA_W
ClientSampleId :
VIBLK17

Manual Integrations
APPROVED

MMDadoda
3/30/2019 12:37:49 PM

Quant Time: Mar 27 09:28:38 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM031419S.M
Quant Title : VOC Analysis
QLast Update : Wed Mar 27 08:01:36 2019
Response via : Initial Calibration









#13
Acetone
Concen: 2.008 ug/L m
RT: 4.13 min Scan# 365
Delta R.T. 0.01 min
Lab File: VW009490.D
Acq: 27 Mar 2019 03:43

Instrument :
MSVOA_W
ClientSampled :
VIBLK17

Tot Ion	Ratio	Lower	Upper
43	100		
58	10.6	0.0	56.2

Manual Integrations
APPROVED

MMDadoda
3/30/2019 12:37:49 PM



#16
Methylene chloride
Concen: 2.007 ug/L m
RT: 4.92 min Scan# 494
Delta R.T. 0.00 min
Lab File: VW009490.D
Acq: 27 Mar 2019 03:43

Tgt Ion	Ratio	Lower	Upper
84	100		
49	132.0	96.0	178.2
86	49.6	44.4	82.5

