

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW032719\
 Data File : VW009523.D
 Acq On : 28 Mar 2019 02:22
 Operator : SY/VA
 Sample : K2068-19
 Misc : 5.12G/10ML/MSVOA W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 BF5J7

Manual Integrations
 APPROVED

apatel
 4/12/2019 8:31:39 AM

Quant Time: Mar 28 08:36:03 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM031419S.M
 Quant Title : VOC Analysis
 QLast Update : Thu Mar 28 07:41:43 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.84	114	279589	25.00	ug/L	0.00
28) Chlorobenzene-d5	11.63	117	217568	25.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.56	152	58163	25.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.36	65	64425	19.28	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	77.12%
7) Chloroethane-d5	2.89	69	79082	20.23	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	80.92%
10) 1,1-Dichloroethene-d2	4.01	63	105429	14.60	ug/L	-0.01
Spiked Amount	25.000	Range	45 - 110	Recovery	=	58.40%
20) 2-Butanone-d5	7.08	46	12100	8.76	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	17.52%#
24) Chloroform-d	7.64	84	199858	26.19	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	104.76%
26) 1,2-Dichloroethane-d4	8.30	65	124696m	27.13	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	108.52%
29) Benzene-d6	8.27	84	306947	25.58	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	102.32%
33) 1,2-Dichloropropane-d6	9.27	67	113943	31.62	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	126.48%#
37) Toluene-d8	10.32	98	225937	20.36	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	81.44%
38) trans-1,3-Dichloropropene-	10.57	79	20423	11.28	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	45.12%
39) 2-Hexanone-d5	10.93	63	9672	11.08	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	22.16%
48) 1,1,2,2-Tetrachloroethane-	12.69	84	100315	28.26	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	113.04%
61) 1,2-Dichlorobenzene-d4	13.86	152	50840	22.80	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	91.20%

Target Compounds				Ovalue
13) Acetone	4.12	43	4171m	3.681 ug/L
16) Methylene chloride	4.91	84	23267m	5.707 ug/L

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW032719\
Data File : VW009523.D
Acq On : 28 Mar 2019 02:22
Operator : SY/VA
Sample : K2068-19
Misc : 5.12G/10ML/MSVOA W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BF5J7

Manual Integrations
APPROVED

apatel
4/12/2019 8:31:39 AM

Quant Time: Mar 28 08:36:03 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM031419S.M
Quant Title : VOC Analysis
QLast Update : Thu Mar 28 07:41:43 2019
Response via : Initial Calibration

