

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW032919\
 Data File : VW009571.D
 Acq On : 29 Mar 2019 14:53
 Operator : SY/VA
 Sample : VW0329SBL01
 Misc : 5.00G/5ML/MSVOA W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VW0329SBL01

Quant Time: Mar 29 17:07:32 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W032919S.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 29 13:17:27 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	266545	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.84	114	433116	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	388799	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.56	152	185078	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.30	65	169279	55.54	ug/l	0.00
Spiked Amount 50.000			Recovery	=	111.08%	
35) Dibromofluoromethane	7.88	113	146181	53.74	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.48%	
50) Toluene-d8	10.32	98	530552	52.77	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.54%	
62) 4-Bromofluorobenzene	12.62	95	180925	48.23	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.46%	
Target Compounds						
20) Methylene Chloride	4.91	84	15753	2.147	ug/l #	Qvalue 83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW032919\
Data File : VW009571.D
Acq On : 29 Mar 2019 14:53
Operator : SY/VA
Sample : VW0329SBL01
Misc : 5.00G/5ML/MSVOA W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VW0329SBL01

Quant Time: Mar 29 17:07:32 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W032919S.M
Quant Title : SW846 8260
QLast Update : Fri Mar 29 13:17:27 2019
Response via : Initial Calibration

