

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW032921\
 Data File : VW018479.D
 Acq On : 29 Mar 2021 12:29
 Operator : SY/VA
 Sample : M1816-22
 Misc : 5.44G/10ML/MSVOA_W/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 BG4S8

Quant Time: Mar 30 03:11:56 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SOM2WLM031921S.M
 Quant Title : VOC Analysis
 QLast Update : Tue Mar 30 03:10:03 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	437426	25.00	ug/L	0.00
28) Chlorobenzene-d5	11.634	117	350035	25.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.554	152	120838	25.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.355	65	134900	22.16	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	88.64%
7) Chloroethane-d5	2.892	69	119906	22.50	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	90.00%
10) 1,1-Dichloroethene-d2	4.019	63	163011	15.51	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	62.04%
20) 2-Butanone-d5	7.074	46	64133	47.32	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	94.64%
24) Chloroform-d	7.653	84	240693	21.60	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	86.40%
26) 1,2-Dichloroethane-d4	8.305	65	144221	22.17	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	88.68%
29) Benzene-d6	8.275	84	461053	24.78	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	99.12%
33) 1,2-Dichloropropane-d6	9.275	67	133704	24.91	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	99.64%
37) Toluene-d8	10.323	98	403655	22.53	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	90.12%
38) trans-1,3-Dichloroprop...	10.579	79	51572	18.63	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	74.52%
39) 2-Hexanone-d5	10.921	63	51506	49.03	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	98.06%
48) 1,1,2,2-Tetrachloroeth...	12.689	84	110029	22.92	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	91.68%
61) 1,2-Dichlorobenzene-d4	13.853	152	90878	21.15	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	84.60%
Target Compounds						Qvalue
13) Acetone	4.117	43	9067	9.31	ug/L	95
16) Methylene chloride	4.922	84	8282	1.15	ug/L	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW032921\
Data File : VW018479.D
Acq On : 29 Mar 2021 12:29
Operator : SY/VA
Sample : M1816-22
Misc : 5.44G/10ML/MSVOA_W/SOIL
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BG4S8

Quant Time: Mar 30 03:11:56 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SOM2WLM031921S.M
Quant Title : VOC Analysis
QLast Update : Tue Mar 30 03:10:03 2021
Response via : Initial Calibration

