

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW032921\
 Data File : VW018474.D
 Acq On : 29 Mar 2021 09:33
 Operator : SY/VA
 Sample : VW0329SBL01
 Misc : 5.00G/10ML/MSVOA_W/SOIL
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VBLK14

Quant Time: Mar 30 03:10:34 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SOM2WLM031921S.M
 Quant Title : VOC Analysis
 QLast Update : Tue Mar 30 03:10:03 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Difluorobenzene	8.841	114	547261	25.00	ug/L	0.00
28) Chlorobenzene-d5	11.634	117	497600	25.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.560	152	239924	25.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.355	65	143151	18.80	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	75.20%
7) Chloroethane-d5	2.891	69	128120	19.22	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	76.88%
10) 1,1-Dichloroethene-d2	4.019	63	192820	14.66	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	58.64%
20) 2-Butanone-d5	7.073	46	80099	47.24	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	94.48%
24) Chloroform-d	7.646	84	312302	22.40	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	89.60%
26) 1,2-Dichloroethane-d4	8.305	65	185228	22.76	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	91.04%
29) Benzene-d6	8.274	84	590850	22.34	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	89.36%
33) 1,2-Dichloropropane-d6	9.274	67	175748	23.03	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	92.12%
37) Toluene-d8	10.323	98	558471	21.93	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	87.72%
38) trans-1,3-Dichloroprop...	10.579	79	84195	21.39	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	85.56%
39) 2-Hexanone-d5	10.920	63	68294	45.73	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	91.46%
48) 1,1,2,2-Tetrachloroeth...	12.694	84	152921	22.40	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	89.60%
61) 1,2-Dichlorobenzene-d4	13.853	152	192053	22.52	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	90.08%
Target Compounds						
16) Methylene chloride	4.915	84	12804	1.43	ug/L	Qvalue 84


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW032921\
Data File : VW018474.D
Acq On : 29 Mar 2021 09:33
Operator : SY/VA
Sample : VW0329SBL01
Misc : 5.00G/10ML/MSVOA_W/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VBLK14

Quant Time: Mar 30 03:10:34 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SOM2WLM031921S.M
Quant Title : VOC Analysis
QLast Update : Tue Mar 30 03:10:03 2021
Response via : Initial Calibration





#16
Methylene chloride
Concen: 1.43 ug/L
RT: 4.915 min Scan# 494
Delta R.T. -0.006 min
Lab File: VW018474.D
Acq: 29 Mar 2021 09:33

Instrument :
MSVOA_W
ClientSampleId :
VBLK14

Tgt Ion:	84	Resp:	12804
Ion	Ratio	Lower	Upper
84	100		
49	86.7	74.4	138.2
86	56.8	46.1	85.7

