

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_W\DATA\VW033019\
 Data File : VW009596.D
 Acq On : 30 Mar 2019 17:56
 Operator : SY/VA
 Sample : K2167-11
 Misc : 6.48G/10ML/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 BF112

Quant Time: Mar 30 22:07:59 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM031419S.M
 Quant Title : VOC Analysis
 QLast Update : Sat Mar 30 21:16:55 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.84	114	423061	25.00	ug/L	0.00
28) Chlorobenzene-d5	11.63	117	408399	25.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.56	152	182903	25.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.36	65	147075	29.09	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	116.36%
7) Chloroethane-d5	2.89	69	143193	24.21	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	96.84%
10) 1,1-Dichloroethene-d2	4.01	63	256767	23.50	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	94.00%
20) 2-Butanone-d5	7.08	46	90514	43.29	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	86.58%
24) Chloroform-d	7.64	84	305632	26.47	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	105.88%
26) 1,2-Dichloroethane-d4	8.31	65	190552	27.40	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	109.60%
29) Benzene-d6	8.27	84	575733	25.56	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	102.24%
33) 1,2-Dichloropropane-d6	9.27	67	175560	25.95	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	103.80%
37) Toluene-d8	10.32	98	512448	24.61	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	98.44%
38) trans-1,3-Dichloropropene-	10.58	79	74951	22.06	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	88.24%
39) 2-Hexanone-d5	10.93	63	70283	42.90	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	85.80%
48) 1,1,2,2-Tetrachloroethane-	12.69	84	155631	23.36	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	93.44%
61) 1,2-Dichlorobenzene-d4	13.86	152	187304	26.71	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	106.84%

Target Compounds

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
19) 1,1-Dichloroethane	6.21	63	22464	2.249	ug/L #	93
22) cis-1,2-Dichloroethene	7.16	96	47062	7.932	ug/L	88
35) Trichloroethene	9.09	95	41947	7.042	ug/L	97
47) Tetrachloroethene	10.86	164	2575757	511.793	ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_W\DATA\VW033019\
Data File : VW009596.D
Acq On : 30 Mar 2019 17:56
Operator : SY/VA
Sample : K2167-11
Misc : 6.48G/10ML/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BF112

Quant Time: Mar 30 22:07:59 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM031419S.M
Quant Title : VOC Analysis
QLast Update : Sat Mar 30 21:16:55 2019
Response via : Initial Calibration

