

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW033023\
 Data File : VW025512.D
 Acq On : 30 Mar 2023 18:59
 Operator : SY/MD
 Sample : 02130-14
 Misc : 6.06g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 C11A3

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 03/31/2023
 Supervised By :Mahesh Dadoda 04/03/2023

Quant Time: Mar 31 02:25:44 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM030623SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Mar 31 02:21:49 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	540043	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	279431	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	38396	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.363	65	118585	13.207	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	52.840%
7) Chloroethane-d5	2.899	69	144400	22.350	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	89.400%
11) 1,1-Dichloroethene-d2	4.021	63	261211	15.061	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	60.240%
21) 2-Butanone-d5	7.081	46	185438	84.135	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	168.280%#
24) Chloroform-d	7.648	84	504172	32.384	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	129.520%
26) 1,2-Dichloroethane-d4	8.307	65	235264	28.154	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	112.600%
32) Benzene-d6	8.276	84	930585	48.739	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	194.960%#
36) 1,2-Dichloropropane-d6	9.270	67	343834	58.210	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	232.840%#
41) Toluene-d8	10.325	98	615358	36.554	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	146.200%#
43) trans-1,3-Dichloroprop...	10.581	79	39629	16.849	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	67.400%
47) 2-Hexanone-d5	10.922	63	74885	88.165	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	176.340%#
56) 1,1,2,2-Tetrachloroeth...	12.690	84	225783	52.175	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	208.720%#
66) 1,2-Dichlorobenzene-d4	13.848	152	64083	44.881	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	179.520%#
Target Compounds					Qvalue	
13) Acetone	4.124	43	53013	25.021	ug/L	98
16) Methylene chloride	4.923	84	25991m	3.456	ug/L	
22) 2-Butanone	7.179	43	16448	6.001	ug/L	92
46) Tetrachloroethene	10.861	164	5310	1.592	ug/L	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW033023\
Data File : VW025512.D
Acq On : 30 Mar 2023 18:59
Operator : SY/MD
Sample : 02130-14
Misc : 6.06g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
C11A3

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 03/31/2023
Supervised By :Mahesh Dadoda 04/03/2023

Quant Time: Mar 31 02:25:44 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM030623SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Mar 31 02:21:49 2023
Response via : Initial Calibration

