

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW033023\
 Data File : VW025503.D
 Acq On : 30 Mar 2023 15:24
 Operator : SY/MD
 Sample : 02130-11
 Misc : 3.93g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 C11A0

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 03/31/2023
 Supervised By :Mahesh Dadoda 04/03/2023

Quant Time: Mar 31 02:23:48 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM030623SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Mar 31 02:21:49 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.849	114	140629	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	82849	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	14244	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.363	65	38055	16.275	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	65.120%
7) Chloroethane-d5	2.893	69	45483	27.034	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	108.120%
11) 1,1-Dichloroethene-d2	4.021	63	82483	18.264	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	73.040%
21) 2-Butanone-d5	7.082	46	34015	59.266	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	118.540%
24) Chloroform-d	7.648	84	125673	30.999	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	124.000%
26) 1,2-Dichloroethane-d4	8.307	65	67423m	30.985	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	123.920%
32) Benzene-d6	8.276	84	233669	41.277	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	165.120%#
36) 1,2-Dichloropropane-d6	9.276	67	81824	46.722	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	186.880%#
41) Toluene-d8	10.325	98	154914	31.037	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	124.160%
43) trans-1,3-Dichloroprop...	10.581	79	17974	25.774	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	103.080%
47) 2-Hexanone-d5	10.928	63	14409	57.217	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	114.440%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	52000	40.529	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	162.120%#
66) 1,2-Dichlorobenzene-d4	13.849	152	14652	27.661	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	110.640%
Target Compounds						Qvalue
16) Methylene chloride	4.911	84	9243m	4.720	ug/L	
46) Tetrachloroethene	10.861	164	3438	3.476	ug/L	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW033023\
Data File : VW025503.D
Acq On : 30 Mar 2023 15:24
Operator : SY/MD
Sample : 02130-11
Misc : 3.93g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1


Instrument :
MSVOA_W
ClientSampleId :
C11A0

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 03/31/2023
Supervised By :Mahesh Dadoda 04/03/2023


Quant Time: Mar 31 02:23:48 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM030623SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Mar 31 02:21:49 2023
Response via : Initial Calibration





#16
Methylene chloride
Concen: 4.720 ug/L m
RT: 4.911 min Scan# 497
Delta R.T. -0.006 min
Lab File: VW025503.D
Acq: 30 Mar 2023 15:24




Instrument :
MSVOA_W
ClientSampleId :
C11A0




Tgt Ion: 84 Resp: 924
Ion Ratio Lower Upper
84 100
86 61.3 45.1 83.7
49 121.6 94.7 175.9

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 03/31/2023
Supervised By :Mahesh Dadoda 04/03/2023



#46
Tetrachloroethene
Concen: 3.476 ug/L
RT: 10.861 min Scan# 1472
Delta R.T. 0.000 min
Lab File: VW025503.D
Acq: 30 Mar 2023 15:24



Tgt Ion:164 Resp: 3438
Ion Ratio Lower Upper
164 100
129 91.6 67.7 125.7
131 85.7 67.0 124.4
166 134.9 86.5 160.7

