

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW033123\
 Data File : VW025532.D
 Acq On : 31 Mar 2023 15:01
 Operator : SY/MD
 Sample : 02151-01
 Misc : 5.89g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 C11A4

Quant Time: Apr 01 00:38:57 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM030623SMA.M
 Quant Title : SFAM01.0
 QLast Update : Sat Apr 01 00:35:04 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	1025254	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	892136	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	397352	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.363	65	165430	9.704	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	38.800%
7) Chloroethane-d5	2.899	69	176836	14.417	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	57.680%
11) 1,1-Dichloroethene-d2	4.021	63	385532	11.709	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	46.840%
21) 2-Butanone-d5	7.075	46	193132	46.156	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	92.320%
24) Chloroform-d	7.648	84	614074	20.776	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	83.120%
26) 1,2-Dichloroethane-d4	8.307	65	344834	21.737	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	86.960%
32) Benzene-d6	8.276	84	1251576	20.531	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	82.120%
36) 1,2-Dichloropropane-d6	9.270	67	412100	21.852	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	87.400%
41) Toluene-d8	10.319	98	1091908	20.316	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	81.280%
43) trans-1,3-Dichloroprop...	10.575	79	152358	20.289	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	81.160%
47) 2-Hexanone-d5	10.922	63	128450	47.368	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	94.740%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	312870	22.645	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	90.600%
66) 1,2-Dichlorobenzene-d4	13.849	152	337738	22.856	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	91.440%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW033123\
Data File : VW025532.D
Acq On : 31 Mar 2023 15:01
Operator : SY/MD
Sample : 02151-01
Misc : 5.89g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
C11A4

Quant Time: Apr 01 00:38:57 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM030623SMA.M
Quant Title : SFAM01.0
QLast Update : Sat Apr 01 00:35:04 2023
Response via : Initial Calibration

