

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW040319\
 Data File : VW009683.D
 Acq On : 04 Apr 2019 04:05
 Operator : SY/VA
 Sample : K2188-05
 Misc : 6.46G/10ML/MSVOA W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 BF661

Quant Time: Apr 04 06:44:06 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM040319S.M
 Quant Title : VOC Analysis
 QLast Update : Thu Apr 04 02:03:18 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.84	114	384819	25.00	ug/L	0.00
28) Chlorobenzene-d5	11.63	117	318604	25.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.56	152	105892	25.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.36	65	146479	29.09	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	116.36%
7) Chloroethane-d5	2.89	69	122761	26.61	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	106.44%
10) 1,1-Dichloroethene-d2	4.01	63	221883	20.49	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	81.96%
20) 2-Butanone-d5	7.07	46	111679	68.31	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	136.62%#
24) Chloroform-d	7.65	84	277833	26.46	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	105.84%
26) 1,2-Dichloroethane-d4	8.30	65	191093	29.64	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	118.56%
29) Benzene-d6	8.27	84	539655	32.27	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	129.08%
33) 1,2-Dichloropropane-d6	9.27	67	168459	32.61	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	130.44%#
37) Toluene-d8	10.32	98	446683	28.89	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	115.56%
38) trans-1,3-Dichloropropene-	10.58	79	63042	25.59	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	102.36%
39) 2-Hexanone-d5	10.93	63	75742	77.79	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	155.58%#
48) 1,1,1,2,2-Tetrachloroethane-	12.69	84	135194	28.15	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	112.60%
61) 1,2-Dichlorobenzene-d4	13.85	152	107095	28.82	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	115.28%

Target Compounds

					Ovalue
13) Acetone	4.12	43	37332	21.310 ug/L	98
14) Carbon disulfide	4.38	76	19770	1.244 ug/L	99
16) Methylene chloride	4.91	84	30227	4.454 ug/L	97
21) 2-Butanone	7.16	43	9873	4.431 ug/L	98
53) Ethylbenzene	11.73	91	14113	0.610 ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW040319\
Data File : VW009683.D
Acq On : 04 Apr 2019 04:05
Operator : SY/VA
Sample : K2188-05
Misc : 6.46G/10ML/MSVOA W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BF661

Quant Time: Apr 04 06:44:06 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM040319S.M
Quant Title : VOC Analysis
QLast Update : Thu Apr 04 02:03:18 2019
Response via : Initial Calibration

