

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW040424\
 Data File : VW028306.D
 Acq On : 04 Apr 2024 11:08
 Operator : SY/MD
 Sample : P1965-04RE
 Misc : 4.53g/10mL/MSVOA_W/SOIL/C
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :

Quant Time: Apr 05 01:00:35 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM031424SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Apr 05 00:59:55 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	308591	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	277895	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.555	152	132517	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.362	65	49611	13.740	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	54.960%
7) Chloroethane-d5	2.905	69	39340	15.595	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	62.400%
11) 1,1-Dichloroethene-d2	4.021	65	31277	13.660	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	54.640%
21) 2-Butanone-d5	7.081	46	34516	32.412	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	64.820%
24) Chloroform-d	7.648	84	138745	15.247	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	61.000%
26) 1,2-Dichloroethane-d4	8.306	65	73059	14.569	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	58.280%#
32) Benzene-d6	8.276	84	269859	15.905	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	63.640%
36) 1,2-Dichloropropane-d6	9.270	67	82591	16.915	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	67.680%#
41) Toluene-d8	10.324	98	237517	15.177	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	60.720%
43) trans-1,3-Dichloroprop...	10.574	79	31673	14.283	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	57.120%
47) 2-Hexanone-d5	10.922	63	26152	32.404	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	64.800%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	61942	17.174	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	68.680%
66) 1,2-Dichlorobenzene-d4	13.848	152	80567	16.690	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	66.760%#

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW040424\
Data File : VW028306.D
Acq On : 04 Apr 2024 11:08
Operator : SY/MD
Sample : P1965-04RE
Misc : 4.53g/10mL/MSVOA_W/SOIL/C
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :

Quant Time: Apr 05 01:00:35 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM031424SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Apr 05 00:59:55 2024
Response via : Initial Calibration

