

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW040424\
 Data File : VW028308.D
 Acq On : 04 Apr 2024 12:41
 Operator : SY/MD
 Sample : P1988-04
 Misc : 5.20g/10mL/MSVOA_W/SOIL/A
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 E07T0

Quant Time: Apr 05 01:01:01 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM031424SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Apr 05 00:59:55 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	459485	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	404805	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	183298	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.362	65	78310	14.566	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	58.280%
7) Chloroethane-d5	2.905	69	61522	16.379	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	65.520%
11) 1,1-Dichloroethene-d2	4.027	65	50506	14.814	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	59.240%
21) 2-Butanone-d5	7.081	46	48412	30.531	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	61.060%
24) Chloroform-d	7.648	84	221023	16.313	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	65.240%
26) 1,2-Dichloroethane-d4	8.306	65	112581	15.078	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	60.320%#
32) Benzene-d6	8.276	84	435204	17.609	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	70.440%
36) 1,2-Dichloropropane-d6	9.270	67	130322	18.323	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	73.280%
41) Toluene-d8	10.324	98	389953	17.106	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	68.440%
43) trans-1,3-Dichloroprop...	10.574	79	49520	15.330	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	61.320%
47) 2-Hexanone-d5	10.922	63	38508	32.755	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	65.500%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	90115	17.152	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	68.600%
66) 1,2-Dichlorobenzene-d4	13.848	152	121552	18.204	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	72.800%#

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW040424\
Data File : VW028308.D
Acq On : 04 Apr 2024 12:41
Operator : SY/MD
Sample : P1988-04
Misc : 5.20g/10mL/MSVOA_W/SOIL/A
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
E07T0

Quant Time: Apr 05 01:01:01 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM031424SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Apr 05 00:59:55 2024
Response via : Initial Calibration

