

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW040424\
 Data File : VW028312.D
 Acq On : 04 Apr 2024 14:16
 Operator : SY/MD
 Sample : P1988-01
 Misc : 4.68g/10mL/MSVOA_W/SOIL/A
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 E07S7

Quant Time: Apr 05 01:02:29 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM031424SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Apr 05 00:59:55 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	379697	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	317653	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	114486	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.363	65	59931	13.490	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	53.960%
7) Chloroethane-d5	2.905	69	47683	15.363	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	61.440%
11) 1,1-Dichloroethene-d2	4.027	65	36657	13.012	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	52.040%
21) 2-Butanone-d5	7.075	46	50945	38.880	ug/L	-0.01
Spiked Amount	50.000	Range	20 - 135	Recovery	=	77.760%
24) Chloroform-d	7.655	84	174584	15.593	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	62.360%
26) 1,2-Dichloroethane-d4	8.301	65	97284	15.767	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	63.080%#
32) Benzene-d6	8.276	84	333730	17.208	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	68.840%
36) 1,2-Dichloropropane-d6	9.270	67	106178	19.024	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	76.080%
41) Toluene-d8	10.325	98	286104	15.994	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	63.960%
43) trans-1,3-Dichloroprop...	10.575	79	39102	15.426	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	61.720%
47) 2-Hexanone-d5	10.922	63	39762	43.101	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	86.200%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	84542	20.506	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	82.040%
66) 1,2-Dichlorobenzene-d4	13.849	152	78023	18.708	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	74.840%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW040424\
Data File : VW028312.D
Acq On : 04 Apr 2024 14:16
Operator : SY/MD
Sample : P1988-01
Misc : 4.68g/10mL/MSVOA_W/SOIL/A
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
E07S7

Quant Time: Apr 05 01:02:29 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM031424SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Apr 05 00:59:55 2024
Response via : Initial Calibration

