

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW040524\
 Data File : VW028329.D
 Acq On : 05 Apr 2024 10:31
 Operator : SY/MD
 Sample : VW0405SBL01
 Misc : 5.00g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VBLK458

Quant Time: Apr 06 00:07:35 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM031424SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Apr 05 23:38:30 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	237908	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	216153	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	108478	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.363	65	42565	15.291	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	61.160%
7) Chloroethane-d5	2.899	69	37650	19.360	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	77.440%
11) 1,1-Dichloroethene-d2	4.015	65	27754	15.723	ug/L	-0.01
Spiked Amount	25.000	Range	45 - 110	Recovery	=	62.880%
21) 2-Butanone-d5	7.075	46	44524	54.231	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	108.460%
24) Chloroform-d	7.648	84	145745	20.775	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	83.120%
26) 1,2-Dichloroethane-d4	8.307	65	80626	20.855	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	83.440%
32) Benzene-d6	8.276	84	276591	20.959	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	83.840%
36) 1,2-Dichloropropane-d6	9.270	67	89733	23.628	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	94.520%
41) Toluene-d8	10.319	98	242512	19.923	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	79.680%
43) trans-1,3-Dichloroprop...	10.575	79	34290	19.879	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	79.520%
47) 2-Hexanone-d5	10.922	63	32370	51.565	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	103.120%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	72199	25.736	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	102.960%
66) 1,2-Dichlorobenzene-d4	13.849	152	85776	21.706	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	86.840%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW040524\
Data File : VW028329.D
Acq On : 05 Apr 2024 10:31
Operator : SY/MD
Sample : VW0405SBL01
Misc : 5.00g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VBLK458

Quant Time: Apr 06 00:07:35 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM031424SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Apr 05 23:38:30 2024
Response via : Initial Calibration

