

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW040524\
 Data File : VW028334.D
 Acq On : 05 Apr 2024 12:58
 Operator : SY/MD
 Sample : P1993-15
 Misc : 5.27g/10mL/MSVOA_W/SOIL/A
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 BH2L6

Quant Time: Apr 05 23:44:20 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM031424SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Apr 05 23:38:30 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	438081	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	403802	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	191640	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.363	65	74397	14.514	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	58.040%
7) Chloroethane-d5	2.899	69	58789	16.416	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	65.680%
11) 1,1-Dichloroethene-d2	4.027	65	48230	14.838	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	59.360%
21) 2-Butanone-d5	7.075	46	64791	42.857	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	85.720%
24) Chloroform-d	7.648	84	221685	17.161	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	68.640%
26) 1,2-Dichloroethane-d4	8.307	65	120190	16.883	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	67.520%#
32) Benzene-d6	8.276	84	432923	17.560	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	70.240%
36) 1,2-Dichloropropane-d6	9.276	67	134512	18.959	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	75.840%
41) Toluene-d8	10.319	98	386817	17.010	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	68.040%
43) trans-1,3-Dichloroprop...	10.575	79	54060	16.777	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	67.120%
47) 2-Hexanone-d5	10.922	63	50342	42.927	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	85.860%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	108442	20.692	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	82.760%
66) 1,2-Dichlorobenzene-d4	13.848	152	140059	20.062	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	80.240%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW040524\
Data File : VW028334.D
Acq On : 05 Apr 2024 12:58
Operator : SY/MD
Sample : P1993-15
Misc : 5.27g/10mL/MSVOA_W/SOIL/A
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BH2L6

Quant Time: Apr 05 23:44:20 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM031424SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Apr 05 23:38:30 2024
Response via : Initial Calibration

