

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW040619\
 Data File : VW009797.D
 Acq On : 07 Apr 2019 03:58
 Operator : SY/VA
 Sample : K2217-20
 Misc : 5.38G/10ML/MSVOA W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 BF610

Manual Integrations
 APPROVED

MMDadoda
 4/11/2019 11:43:27 AM

Quant Time: Apr 08 00:37:31 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM040319S.M
 Quant Title : VOC Analysis
 QLast Update : Sun Apr 07 05:36:41 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.84	114	372710	25.00	ug/L	0.00
28) Chlorobenzene-d5	11.63	117	354136	25.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.56	152	131590	25.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.36	65	122130	25.04	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	100.16%
7) Chloroethane-d5	2.89	69	120045	26.87	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	107.48%
10) 1,1-Dichloroethene-d2	4.01	63	219356	20.91	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	83.64%
20) 2-Butanone-d5	7.07	46	100380	63.39	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	126.78%
24) Chloroform-d	7.64	84	309950	30.48	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	121.92%
26) 1,2-Dichloroethane-d4	8.30	65	204534	32.76	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	131.04%#
29) Benzene-d6	8.27	84	571873	30.76	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	123.04%
33) 1,2-Dichloropropane-d6	9.27	67	185194	32.25	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	129.00%#
37) Toluene-d8	10.32	98	502183	29.22	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	116.88%
38) trans-1,3-Dichloropropene-	10.58	79	71076	25.95	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	103.80%
39) 2-Hexanone-d5	10.93	63	60698	56.08	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	112.16%
48) 1,1,2,2-Tetrachloroethane-	12.69	84	161702	30.29	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	121.16%#
61) 1,2-Dichlorobenzene-d4	13.85	152	156085	33.80	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	135.20%#

Target Compounds					Ovalue
13) Acetone	4.13	43	13069m	7.703	ug/L
16) Methylene chloride	4.91	84	11482m	1.747	ug/L
44) Toluene	10.38	91	7564	0.335	ug/L

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW040619\
Data File : VW009797.D
Acq On : 07 Apr 2019 03:58
Operator : SY/VA
Sample : K2217-20
Misc : 5.38G/10ML/MSVOA W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BF610

Manual Integrations
APPROVED

MMDadoda
4/11/2019 11:43:27 AM

Quant Time: Apr 08 00:37:31 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM040319S.M
Quant Title : VOC Analysis
QLast Update : Sun Apr 07 05:36:41 2019
Response via : Initial Calibration

