




Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_W\DATA\VW040619\
Data File : VW009768.D
Acq On : 06 Apr 2019 13:12
Operator : SY/VA
Sample : K2277-12
Misc : 4.82G/10ML/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BF175

Manual Integrations
APPROVED

MMDadoda
4/11/2019 11:43:02 AM

Quant Time: Apr 07 05:07:15 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM040319S.M
Quant Title : VOC Analysis
QLast Update : Sun Apr 07 03:56:20 2019
Response via : Initial Calibration



TIC: VW009768.D

(16) Methylene chloride (T)

4.909min (-0.006) 0.05ug/L

response 481

Ion	Exp%	Act%
84.00	100	100
49.00	140.40	115.47
86.00	66.30	60.21
0.00	0.00	0.00




Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_W\DATA\VW040619\
Data File : VW009768.D
Acq On : 06 Apr 2019 13:12
Operator : SY/VA
Sample : K2277-12
Misc : 4.82G/10ML/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BF175

Manual Integrations
APPROVED

MMDadoda
4/11/2019 11:43:02 AM

Quant Time: Apr 07 05:07:15 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM040319S.M
Quant Title : VOC Analysis
QLast Update : Sun Apr 07 03:56:20 2019
Response via : Initial Calibration



TIC: VW009768.D

(16) Methylene chloride (T)
4.909min (-0.006) 0.69ug/L m
response 6233

Ion	Exp%	Act%
84.00	100	100
49.00	140.40	115.47
86.00	66.30	60.21
0.00	0.00	0.00

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_W\DATA\VW040619\
 Data File : VW009768.D
 Acq On : 06 Apr 2019 13:12
 Operator : SY/VA
 Sample : K2277-12
 Misc : 4.82G/10ML/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 BF175

Manual Integrations
 APPROVED

MMDadoda
 4/11/2019 11:43:02 AM

Quant Time: Apr 07 05:13:01 2019

Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM040319S.M

Quant Title : VOC Analysis

QLast Update : Sun Apr 07 03:56:20 2019

Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.84	114	512223	25.00	ug/L	0.00
28) Chlorobenzene-d5	11.63	117	483561	25.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.56	152	221408	25.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.35	65	164030	24.47	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	97.88%
7) Chloroethane-d5	2.90	69	148830	24.24	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	96.96%
10) 1,1-Dichloroethene-d2	4.01	63	286041	19.84	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	79.36%
20) 2-Butanone-d5	7.07	46	97578	44.84	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	89.68%
24) Chloroform-d	7.64	84	365581	26.16	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	104.64%
26) 1,2-Dichloroethane-d4	8.31	65	221875	25.86	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	103.44%
29) Benzene-d6	8.27	84	706826	27.84	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	111.36%
33) 1,2-Dichloropropane-d6	9.27	67	213770	27.26	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	109.04%
37) Toluene-d8	10.32	98	643148	27.40	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	109.60%
38) trans-1,3-Dichloropropene-	10.58	79	89878	24.04	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	96.16%
39) 2-Hexanone-d5	10.93	63	68826	46.57	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	93.14%
48) 1,1,2,2-Tetrachloroethane-	12.69	84	165130	22.65	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	90.60%
61) 1,2-Dichlorobenzene-d4	13.85	152	218013	28.06	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	112.24%

Target Compounds

					Qvalue
13) Acetone	4.12	43	6129	2.628 ug/L	89
22) cis-1,2-Dichloroethene	7.17	96	28347	3.746 ug/L	98
35) Trichloroethene	9.09	95	153645	20.180 ug/L	100
47) Tetrachloroethene	10.86	164	121527	20.266 ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_W\DATA\VW040619\
Data File : VW009768.D
Acq On : 06 Apr 2019 13:12
Operator : SY/VA
Sample : K2277-12
Misc : 4.82G/10ML/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BF175

Manual Integrations
APPROVED

MMDadoda
4/11/2019 11:43:02 AM

Quant Time: Apr 07 05:13:01 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM040319S.M
Quant Title : VOC Analysis
QLast Update : Sun Apr 07 03:56:20 2019
Response via : Initial Calibration

