

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_W\DATA\VW040819\
 Data File : VW009810.D
 Acq On : 08 Apr 2019 19:53
 Operator : SY/VA
 Sample : K2259-04
 Misc : 4.03G/10ML/MSVOA_W/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 ETJ43RE

Quant Time: Apr 09 02:40:26 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM040319S.M
 Quant Title : VOC Analysis
 QLast Update : Tue Apr 09 02:37:30 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.84	114	345906	25.00	ug/L	0.00
28) Chlorobenzene-d5	11.63	117	292178	25.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.56	152	93413	25.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.35	65	116530	25.74	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	102.96%
7) Chloroethane-d5	2.89	69	108587	26.18	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	104.72%
10) 1,1-Dichloroethene-d2	4.01	63	203619	20.92	ug/L	-0.01
Spiked Amount	25.000	Range	45 - 110	Recovery	=	83.68%
20) 2-Butanone-d5	7.08	46	137798	93.76	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	187.52%#
24) Chloroform-d	7.64	84	291466	30.88	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	123.52%
26) 1,2-Dichloroethane-d4	8.30	65	194902	33.63	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	134.52%#
29) Benzene-d6	8.27	84	517333	33.73	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	134.92%
33) 1,2-Dichloropropane-d6	9.27	67	169235	35.72	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	142.88%#
37) Toluene-d8	10.32	98	420980	29.69	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	118.76%
38) trans-1,3-Dichloropropene-	10.58	79	66651	29.50	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	118.00%
39) 2-Hexanone-d5	10.93	63	82136	91.99	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	183.98%#
48) 1,1,2,2-Tetrachloroethane-	12.69	84	157913	35.85	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	143.40%#
61) 1,2-Dichlorobenzene-d4	13.85	152	104909	32.01	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	128.04%#

Target Compounds

					Qvalue
13) Acetone	4.12	43	79328	50.377 ug/L	100
16) Methylene chloride	4.92	84	12896	2.114 ug/L	88
21) 2-Butanone	7.17	43	31085	15.522 ug/L	98
42) cis-1,3-Dichloropropene	10.03	75	5713	0.819 ug/L	84
59) 1,2,3-Trichloropropane	13.56	75	12694	3.755 ug/L	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_W\DATA\VW040819\
Data File : VW009810.D
Acq On : 08 Apr 2019 19:53
Operator : SY/VA
Sample : K2259-04
Misc : 4.03G/10ML/MSVOA_W/SOIL
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
ETJ43RE

Quant Time: Apr 09 02:40:26 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM040319S.M
Quant Title : VOC Analysis
QLast Update : Tue Apr 09 02:37:30 2019
Response via : Initial Calibration

