

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW040819\
 Data File : VW009854.D
 Acq On : 09 Apr 2019 23:12
 Operator : SY/VA
 Sample : K2327-11
 Misc : 6.03G/10ML/MSVOA W/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 BF0N5

Manual Integrations
 APPROVED

MMDadoda
 4/11/2019 2:05:46 PM

Quant Time: Apr 10 05:45:58 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM040319S.M
 Quant Title : VOC Analysis
 QLast Update : Wed Apr 10 04:42:28 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.84	114	429844	25.00	ug/L	0.00
28) Chlorobenzene-d5	11.63	117	397901	25.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.56	152	145198	25.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.35	65	119815	21.30	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	85.20%
7) Chloroethane-d5	2.89	69	112759	21.88	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	87.52%
10) 1,1-Dichloroethene-d2	4.01	63	222765	18.42	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	73.68%
20) 2-Butanone-d5	7.08	46	128348	70.28	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	140.56%#
24) Chloroform-d	7.65	84	316011	26.94	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	107.76%
26) 1,2-Dichloroethane-d4	8.31	65	209093m	29.04	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	116.16%
29) Benzene-d6	8.27	84	570359	27.31	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	109.24%
33) 1,2-Dichloropropane-d6	9.27	67	182720	28.32	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	113.28%
37) Toluene-d8	10.32	98	489958	25.37	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	101.48%
38) trans-1,3-Dichloropropene-	10.58	79	73280	23.82	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	95.28%
39) 2-Hexanone-d5	10.93	63	80230	65.98	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	131.96%
48) 1,1,2,2-Tetrachloroethane-	12.69	84	180094	30.03	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	120.12%#
61) 1,2-Dichlorobenzene-d4	13.86	152	150980	29.63	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	118.52%

Target Compounds

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
13) Acetone	4.12	43	58351	29.820	ug/L	98
14) Carbon disulfide	4.37	76	8843m	0.498	ug/L	
16) Methylene chloride	4.92	84	8670m	1.144	ug/L	
21) 2-Butanone	7.17	43	13791	5.542	ug/L	93
22) cis-1,2-Dichloroethene	7.17	96	32341	5.093	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW040819\
Data File : VW009854.D
Acq On : 09 Apr 2019 23:12
Operator : SY/VA
Sample : K2327-11
Misc : 6.03G/10ML/MSVOA W/SOIL
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BF0N5

Manual Integrations
APPROVED

MMDadoda
4/11/2019 2:05:46 PM

Quant Time: Apr 10 05:45:58 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM040319S.M
Quant Title : VOC Analysis
QLast Update : Wed Apr 10 04:42:28 2019
Response via : Initial Calibration

