

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW041019\
 Data File : VW009883.D
 Acq On : 10 Apr 2019 16:10
 Operator : SY/VA
 Sample : VW0410SBL01
 Misc : 5.00G/10ML/MSVOA W/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VBLK02

Manual Integrations
 APPROVED

MMDadoda
 4/15/2019 12:33:12 AM

Quant Time: Apr 11 03:56:06 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM040319S.M
 Quant Title : VOC Analysis
 QLast Update : Thu Apr 11 03:44:42 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.84	114	480425	25.00	ug/L	0.00
28) Chlorobenzene-d5	11.63	117	458216	25.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.56	152	207203	25.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.36	65	165781	26.37	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	105.48%
7) Chloroethane-d5	2.90	69	129672	22.51	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	90.04%
10) 1,1-Dichloroethene-d2	4.01	63	277846	20.55	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	82.20%
20) 2-Butanone-d5	7.07	46	121425	59.49	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	118.98%
24) Chloroform-d	7.65	84	320565	24.45	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	97.80%
26) 1,2-Dichloroethane-d4	8.31	65	205971	25.59	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	102.36%
29) Benzene-d6	8.27	84	621788	25.85	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	103.40%
33) 1,2-Dichloropropane-d6	9.27	67	186184	25.06	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	100.24%
37) Toluene-d8	10.32	98	550643	24.76	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	99.04%
38) trans-1,3-Dichloropropene-	10.57	79	83092	23.45	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	93.80%
39) 2-Hexanone-d5	10.93	63	77535	55.37	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	110.74%
48) 1,1,2,2-Tetrachloroethane-	12.69	84	173241	25.08	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	100.32%
61) 1,2-Dichlorobenzene-d4	13.86	152	187212	25.75	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	103.00%

Target Compounds					Ovalue
13) Acetone	4.12	43	7869	3.598 ug/L	91
16) Methylene chloride	4.92	84	25131m	2.966 ug/L	

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW041019\
Data File : VW009883.D
Acq On : 10 Apr 2019 16:10
Operator : SY/VA
Sample : VW0410SBL01
Misc : 5.00G/10ML/MSVOA W/SOIL
ALS Vial : 2 Sample Multiplier: 1


Instrument :
MSVOA_W
ClientSampleId :
VBLK02

Manual Integrations
APPROVED

MMDadoda
4/15/2019 12:33:12 AM

Quant Time: Apr 11 03:56:06 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM040319S.M
Quant Title : VOC Analysis
QLast Update : Thu Apr 11 03:44:42 2019
Response via : Initial Calibration





#13

Acetone

Concen: 3.598 ug/L

RT: 4.12 min Scan# 364

Delta R.T. 0.00 min

Lab File: VW009883.D

Acq: 10 Apr 2019 16:10

Instrument :

MSVOA_W

ClientSampleId :

VBLK02

Tot Ion: 43 Resp: 7869

Ion Ratio Lower Upper

43 100


58 22.1 0.0 53.0

Manual Integrations

APPROVED




MMDadoda

4/15/2019 12:33:12 AM



Abundance Ion 43.00 (42.70 to 43.70): VW

Ion 58.00 (57.70 to 58.70): VW



#16

Methylene chloride

Concen: 2.966 ug/L m

RT: 4.92 min Scan# 494

Delta R.T. 0.00 min

Lab File: VW009883.D

Acq: 10 Apr 2019 16:10


Tgt Ion: 84 Resp: 25131

Ion Ratio Lower Upper

84 100

49 141.1 98.3 182.5

86 63.8 46.4 86.2



Abundance Ion 84.00 (83.70 to 84.70): VW

Ion 49.00 (48.70 to 49.70): VW

Ion 86.00 (85.70 to 86.70): VW

