

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW041020\
 Data File : VW015182.D
 Acq On : 10 Apr 2020 16:08
 Operator : SY/VA
 Sample : L2176-09
 Misc : 5.08G/10ML/MSVOA W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 BG2J5

Quant Time: Apr 11 07:14:10 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM041020S.M
 Quant Title : VOC Analysis
 QLast Update : Fri Apr 10 13:10:27 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.84	114	1200712	25.00	ug/L	0.00
28) Chlorobenzene-d5	11.63	117	1033207	25.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.55	152	442383	25.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.36	65	335046	29.50	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	118.00%
7) Chloroethane-d5	2.89	69	221882	26.65	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	106.60%
10) 1,1-Dichloroethene-d2	4.02	63	543734	18.08	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	72.32%
20) 2-Butanone-d5	7.07	46	156637	40.33	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	80.66%
24) Chloroform-d	7.65	84	637495	22.51	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	90.04%
26) 1,2-Dichloroethane-d4	8.31	65	300116	21.40	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	85.60%
29) Benzene-d6	8.27	84	1445151	25.28	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	101.12%
33) 1,2-Dichloropropane-d6	9.27	67	414117	23.43	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	93.72%
37) Toluene-d8	10.32	98	1313045	24.51	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	98.04%
38) trans-1,3-Dichloropropene-	10.57	79	175364	21.74	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	86.96%
39) 2-Hexanone-d5	10.92	63	133319	42.22	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	84.44%
48) 1,1,2,2-Tetrachloroethane-	12.69	84	269877	20.89	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	83.56%
61) 1,2-Dichlorobenzene-d4	13.85	152	367754	23.20	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	92.80%

Target Compounds

					Ovalue
13) Acetone	4.13	43	55190	22.417 ug/L	97
14) Carbon disulfide	4.39	76	22910	0.423 ug/L	100
54) m,p-Xylene	11.84	106	14210	0.487 ug/L	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW041020\
Data File : VW015182.D
Acq On : 10 Apr 2020 16:08
Operator : SY/VA
Sample : L2176-09
Misc : 5.08G/10ML/MSVOA W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BG2J5

Quant Time: Apr 11 07:14:10 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM041020S.M
Quant Title : VOC Analysis
QLast Update : Fri Apr 10 13:10:27 2020
Response via : Initial Calibration

