

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW041119\
 Data File : VW009907.D
 Acq On : 11 Apr 2019 11:44
 Operator : SY/VA
 Sample : K2254-20
 Misc : 6.04G/10ML/MSVOA W/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 BFC41

Quant Time: Jun 12 09:22:08 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM040319S.M
 Quant Title : VOC Analysis
 QLast Update : Fri Apr 19 08:16:42 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.84	114	500568	25.00	ug/L	0.00
28) Chlorobenzene-d5	11.63	117	459973	25.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.56	152	176752	25.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.36	65	112730	17.21	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	68.84%
7) Chloroethane-d5	2.89	69	113981	18.99	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	75.96%
10) 1,1-Dichloroethene-d2	4.01	63	222852	15.82	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	63.28%
20) 2-Butanone-d5	7.07	46	86646	40.74	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	81.48%
24) Chloroform-d	7.64	84	362251	26.52	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	106.08%
26) 1,2-Dichloroethane-d4	8.31	65	218077	26.01	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	104.04%
29) Benzene-d6	8.27	84	663619	27.48	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	109.92%
33) 1,2-Dichloropropane-d6	9.27	67	213067	28.57	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	114.28%
37) Toluene-d8	10.32	98	561999	25.17	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	100.68%
38) trans-1,3-Dichloropropene-	10.57	79	76681	21.56	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	86.24%
39) 2-Hexanone-d5	10.93	63	48083	34.21	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	68.42%
48) 1,1,2,2-Tetrachloroethane-	12.69	84	144113	20.78	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	83.12%
61) 1,2-Dichlorobenzene-d4	13.86	152	173900	28.04	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	112.16%

Target Compounds					Ovalue
13) Acetone	4.12	43	14878	6.529 ug/L	95
15) Methyl Acetate	4.67	43	4515	1.087 ug/L #	86
16) Methylene chloride	4.91	84	32110	3.637 ug/L	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW041119\
Data File : VW009907.D
Acq On : 11 Apr 2019 11:44
Operator : SY/VA
Sample : K2254-20
Misc : 6.04G/10ML/MSVOA W/SOIL
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BFC41

Quant Time: Jun 12 09:22:08 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM040319S.M
Quant Title : VOC Analysis
QLast Update : Fri Apr 19 08:16:42 2019
Response via : Initial Calibration

