

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW041122\
 Data File : VW022477.D
 Acq On : 11 Apr 2022 12:14
 Operator : SY/VA
 Sample : VW0411SBL01
 Misc : 5.00g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VBLK476

Quant Time: Apr 11 23:57:02 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM040722SMA.M
 Quant Title : SFAM01.0
 QLast Update : Mon Apr 11 23:55:29 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	407636	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.628	117	393452	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	195833	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.355	65	131051	14.542	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	58.160%
7) Chloroethane-d5	2.891	69	84156	14.901	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	59.600%
11) 1,1-Dichloroethene-d2	4.019	63	150223	12.852	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	51.400%
21) 2-Butanone-d5	7.080	46	67800	43.094	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	86.180%
24) Chloroform-d	7.647	84	266784	20.938	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	83.760%
26) 1,2-Dichloroethane-d4	8.305	65	151413	20.799	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	83.200%
32) Benzene-d6	8.275	84	485442	20.194	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	80.760%
36) 1,2-Dichloropropane-d6	9.274	67	148826	21.394	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	85.560%
41) Toluene-d8	10.323	98	459424	19.514	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	78.040%
43) trans-1,3-Dichloroprop...	10.579	79	62370	18.998	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	76.000%
47) 2-Hexanone-d5	10.920	63	53578	41.396	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	82.800%
56) 1,1,2,2-Tetrachloroeth...	12.688	84	127011	21.473	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	85.880%
66) 1,2-Dichlorobenzene-d4	13.853	152	164720	23.747	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	95.000%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW041122\
Data File : VW022477.D
Acq On : 11 Apr 2022 12:14
Operator : SY/VA
Sample : VW0411SBL01
Misc : 5.00g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VBLK476

Quant Time: Apr 11 23:57:02 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM040722SMA.M
Quant Title : SFAM01.0
QLast Update : Mon Apr 11 23:55:29 2022
Response via : Initial Calibration

