

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW041222\
 Data File : VW022500.D
 Acq On : 12 Apr 2022 12:21
 Operator : SY/VA
 Sample : VIBLK419
 Misc : 5.00g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VIBLK419

Quant Time: Apr 12 23:31:26 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM040722SMA.M
 Quant Title : SFAM01.0
 QLast Update : Tue Apr 12 23:30:19 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.841	114	345713	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.627	117	336197	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	170375	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.355	65	136542	17.866	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	71.480%
7) Chloroethane-d5	2.885	69	92854	19.386	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	77.560%
11) 1,1-Dichloroethene-d2	4.013	63	141127	14.237	ug/L	-0.01
Spiked Amount	25.000	Range	45 - 110	Recovery	=	56.960%
21) 2-Butanone-d5	7.079	46	75512	56.593	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	113.180%
24) Chloroform-d	7.652	84	258679	23.939	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	95.760%
26) 1,2-Dichloroethane-d4	8.305	65	152592	24.715	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	98.880%
32) Benzene-d6	8.274	84	475058	23.128	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	92.520%
36) 1,2-Dichloropropane-d6	9.274	67	144744	24.351	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	97.400%
41) Toluene-d8	10.323	98	442989	22.021	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	88.080%
43) trans-1,3-Dichloroprop...	10.579	79	62702	22.351	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	89.400%
47) 2-Hexanone-d5	10.920	63	64830	58.621	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	117.240%
56) 1,1,2,2-Tetrachloroeth...	12.688	84	138441	27.391	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	109.560%
66) 1,2-Dichlorobenzene-d4	13.847	152	166265	27.552	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	110.200%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW041222\
Data File : VW022500.D
Acq On : 12 Apr 2022 12:21
Operator : SY/VA
Sample : VIBLK419
Misc : 5.00g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VIBLK419

Quant Time: Apr 12 23:31:26 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM040722SMA.M
Quant Title : SFAM01.0
QLast Update : Tue Apr 12 23:30:19 2022
Response via : Initial Calibration

