

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW041321\
 Data File : VW018640.D
 Acq On : 13 Apr 2021 14:12
 Operator : SY/VA
 Sample : M2011-04RE
 Misc : 4.82G/10ML/MSVOA_W/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 BG784RE

Quant Time: Apr 14 01:27:21 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SOM2WLM031921S.M
 Quant Title : VOC Analysis
 QLast Update : Wed Apr 14 01:25:39 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	392521	25.00	ug/L	0.00
28) Chlorobenzene-d5	11.628	117	324662	25.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.560	152	101755	25.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.355	65	123677	22.64	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	90.56%
7) Chloroethane-d5	2.892	69	97117	20.31	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	81.24%
10) 1,1-Dichloroethene-d2	4.019	63	129297	13.71	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	54.84%
20) 2-Butanone-d5	7.074	46	67997	55.92	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	111.84%
24) Chloroform-d	7.647	84	214198	21.42	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	85.68%
26) 1,2-Dichloroethane-d4	8.305	65	134966	23.13	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	92.52%
29) Benzene-d6	8.275	84	420660	24.37	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	97.48%
33) 1,2-Dichloropropane-d6	9.274	67	124683	25.04	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	100.16%
37) Toluene-d8	10.323	98	385873	23.22	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	92.88%
38) trans-1,3-Dichloroprop...	10.579	79	43329	16.88	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	67.52%
39) 2-Hexanone-d5	10.921	63	52283	53.66	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	107.32%
48) 1,1,2,2-Tetrachloroeth...	12.689	84	105103	23.60	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	94.40%
61) 1,2-Dichlorobenzene-d4	13.853	152	78732	21.76	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	87.04%
Target Compounds						
					Qvalue	
13) Acetone	4.123	43	26385	30.20	ug/L	96
14) Carbon disulfide	4.391	76	60867	4.10	ug/L	96
16) Methylene chloride	4.916	84	13704	2.13	ug/L	92
44) Toluene	10.384	91	6747	0.32	ug/L	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW041321\
Data File : VW018640.D
Acq On : 13 Apr 2021 14:12
Operator : SY/VA
Sample : M2011-04RE
Misc : 4.82G/10ML/MSVOA_W/SOIL
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BG784RE

Quant Time: Apr 14 01:27:21 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SOM2WLM031921S.M
Quant Title : VOC Analysis
QLast Update : Wed Apr 14 01:25:39 2021
Response via : Initial Calibration

