

m/z	Abundance
36.00	752.0
37.00	3947.0
38.00	3674.0
39.00	1760.0
39.90	192.0
41.10	92.0
42.90	186.0
44.00	858.0
45.10	705.0
46.00	291.0
47.10	810.0
48.10	543.0
49.00	3665.0
50.10	15454.0
51.00	4793.0
52.10	283.0
52.90	50.0
54.90	327.0
56.10	1619.0
57.00	2475.0
58.20	84.0
59.10	198.0
60.00	866.0
61.00	4406.0
62.00	4424.0
63.10	3991.0
64.10	311.0
64.90	179.0
65.90	53.0
67.20	301.0
68.00	8825.0
69.00	8954.0
70.00	896.0
71.90	494.0
73.00	4508.0
74.10	15699.0
75.00	46920.0
76.00	3943.0
76.90	610.0
78.00	388.0
78.90	3532.0
80.00	1033.0
81.00	3715.0
82.00	978.0
82.90	181.0
86.20	207.0
87.00	2942.0
88.00	3162.0
90.90	346.0
92.00	2494.0
93.00	4542.0
94.00	11312.0
95.00	82072.0
96.00	5692.0
97.20	200.0
103.10	152.0
104.00	433.0
104.80	190.0
105.90	436.0
106.90	92.0
109.80	158.0
110.80	138.0
111.70	76.0
112.30	61.0
112.80	59.0
113.30	52.0
115.10	109.0
115.90	448.0
116.90	821.0
117.80	344.0
118.80	465.0
124.90	296.0
126.70	57.0
128.00	391.0
129.10	176.0
129.80	342.0
130.90	121.0
133.00	607.0
133.90	162.0
134.70	215.0
135.10	199.0
136.80	178.0
140.90	1152.0
142.80	890.0
143.90	53.0
145.00	220.0
145.80	156.0
146.90	154.0
147.70	156.0
148.00	160.0
148.80	73.0
149.90	154.0

Instrument :
MSVOA_W
ClientSampleId :
BFB04

152.60	52.0
155.00	179.0
157.00	128.0
157.80	52.0
158.90	122.0
161.10	64.0
163.00	76.0
165.10	163.0
170.40	145.0
171.00	184.0
173.90	63976.0
175.00	4480.0
175.90	66680.0
176.90	4288.0
178.00	120.0
179.00	141.0
191.00	406.0
191.90	194.0
192.90	742.0
193.80	148.0
194.90	62.0
204.90	133.0
207.10	455.0
208.00	58.0
234.80	60.0
248.80	401.0
250.00	308.0
250.60	81.0
265.20	731.0
265.90	328.0
267.00	64.0
278.00	54.0
278.90	126.0
279.60	94.0
280.00	100.0
281.10	7576.0
282.00	2464.0
282.90	1176.0
284.10	320.0
284.90	172.0
286.10	55.0

Instrument :
MSVOA_W
ClientSampleId :
BFB04

m/z	Abundance
36.10	955.0
37.00	4611.0
38.10	4032.0
39.10	1587.0
39.90	270.0
42.90	142.0
44.00	1165.0
45.00	820.0
46.10	179.0
47.10	736.0
47.90	571.0
49.00	4516.0
50.00	17968.0
51.00	5073.0
52.10	380.0
55.10	343.0
56.00	2037.0
57.00	2933.0
58.10	243.0
59.00	321.0
59.90	900.0
61.00	4840.0
62.10	5042.0
63.00	3738.0
64.00	454.0
64.90	294.0
67.00	396.0
68.00	9531.0
69.00	9781.0
70.10	679.0
72.00	478.0
73.00	5653.0
74.00	16576.0
75.00	49136.0
76.00	4476.0
77.10	392.0
77.80	486.0
78.90	4114.0
79.90	1024.0
80.90	3840.0
81.80	985.0
82.70	134.0
83.10	134.0
85.80	63.0
87.00	3509.0
88.00	2885.0
90.80	565.0
92.00	3033.0
93.00	4091.0
94.00	11774.0
95.00	89584.0
96.00	5808.0
97.20	210.0
103.00	122.0
103.90	591.0
104.90	283.0
106.00	502.0
106.80	161.0
109.70	93.0
110.10	154.0
110.80	175.0
112.00	153.0
112.70	132.0
115.10	133.0
116.10	231.0
116.90	872.0
117.80	375.0
118.90	654.0
124.80	398.0
126.80	77.0
127.80	303.0
128.60	71.0
129.90	269.0
131.00	125.0
133.20	975.0
134.10	199.0
134.90	202.0
136.90	134.0
140.20	106.0
140.90	1078.0
142.90	841.0
145.00	236.0
145.80	175.0
146.80	272.0
147.90	210.0
150.00	101.0
153.00	101.0
154.80	277.0
156.90	171.0
157.90	103.0
158.70	122.0
160.90	161.0

Instrument :
MSVOA_W
ClientSampleId :
BFB04

162.90	218.0
165.00	212.0
169.10	94.0
170.80	176.0
172.10	482.0
173.90	62432.0
175.00	4685.0
176.00	52656.0
177.00	3943.0
177.90	335.0
178.80	195.0
188.80	86.0
189.20	89.0
190.90	967.0
192.00	235.0
193.10	1234.0
194.10	289.0
194.80	202.0
195.80	94.0
202.90	162.0
204.10	66.0
205.10	181.0
207.00	1298.0
207.90	189.0
209.00	138.0
220.70	63.0
234.80	115.0
249.00	1032.0
249.90	276.0
251.00	332.0
252.80	110.0
264.90	798.0
265.90	421.0
267.20	265.0
276.70	69.0
278.20	104.0
279.00	220.0
281.00	13236.0
281.90	3293.0
283.00	2124.0
284.10	459.0
284.90	168.0

Instrument :
MSVOA_W
ClientSampleId :
BFB04

m/z	Abundance
36.00	705.0
37.10	3504.0
37.90	2987.0
39.00	1406.0
39.90	244.0
40.80	62.0
41.30	58.0
43.00	249.0
44.00	913.0
45.00	1014.0
46.10	140.0
46.90	513.0
48.00	457.0
49.10	2719.0
50.00	13714.0
51.00	3683.0
51.90	349.0
55.20	160.0
56.10	1302.0
57.00	1865.0
58.10	79.0
59.00	385.0
60.00	679.0
61.00	4077.0
62.00	3702.0
63.10	3188.0
64.00	261.0
65.00	270.0
67.10	373.0
68.00	8112.0
69.00	7306.0
70.00	886.0
70.80	84.0
71.90	436.0
73.00	4521.0
74.00	12968.0
75.00	35672.0
76.00	3376.0
76.90	411.0
78.00	465.0
79.00	2524.0
79.90	812.0
80.90	2721.0
81.80	603.0
83.20	97.0
87.00	2466.0
87.80	2296.0
89.10	68.0
90.80	261.0
92.00	1642.0
93.10	2680.0
94.00	8941.0
95.00	64208.0
96.00	4679.0
97.00	159.0
103.10	319.0
103.80	505.0
105.00	280.0
106.00	245.0
106.80	131.0
110.00	109.0
111.00	92.0
112.10	155.0
112.80	98.0
115.00	153.0
115.80	531.0
116.90	414.0
118.20	483.0
118.90	906.0
125.00	287.0
125.70	163.0
127.90	275.0
128.60	151.0
129.80	280.0
130.80	124.0
133.10	1404.0
133.80	314.0
134.90	235.0
136.80	114.0
140.90	675.0
142.90	801.0
145.10	138.0
146.10	68.0
146.80	297.0
147.90	246.0
149.20	92.0
149.80	84.0
154.90	132.0
156.90	114.0
160.90	297.0
162.90	145.0
165.10	222.0

Instrument :
MSVOA_W
ClientSampleId :
BFB04

171.50	194.0
171.90	234.0
172.60	751.0
173.90	38624.0
175.00	3990.0
175.90	41184.0
176.90	3106.0
177.90	192.0
179.00	382.0
179.90	95.0
188.70	140.0
189.10	174.0
189.90	76.0
191.00	1198.0
193.00	1951.0
194.10	300.0
194.60	136.0
195.10	156.0
195.90	118.0
202.90	125.0
205.10	352.0
206.90	882.0
208.00	220.0
208.80	143.0
220.80	63.0
248.90	1256.0
250.10	223.0
250.80	442.0
252.80	140.0
265.00	1100.0
266.00	462.0
267.20	265.0
278.20	205.0
278.80	206.0
279.50	301.0
280.10	446.0
281.10	15772.0
282.00	4560.0
283.00	2784.0
283.90	426.0
284.90	279.0

Instrument :
MSVOA_W
ClientSampleId :
BFB04

m/z	Abundance
39.90	190.0
44.10	335.0
68.90	145.0



Instrument :
MSVOA_W
ClientSampleId :
BFB04

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW041321\
 Data File : VW018634.D
 Acq On : 13 Apr 2021 10:20
 Operator : SY/VA
 Sample : BFB04
 Misc : 5.00G/10ML/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 BFB04

Integration File: LSCINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SOM2WLM031921S.M
 Title : VOC Analysis
 Last Update : Wed Apr 14 01:25:39 2021



AutoFind: Scans 1756, 1757, 1758; Background Corrected with Scan 1748

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	20.0	15712	PASS
75	95	30	80	55.8	43909	PASS
95	95	100	100	100.0	78621	PASS
96	95	5	9	6.9	5393	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	120	70.0	55011	PASS
175	174	5	9	8.0	4385	PASS
176	174	95	101	97.3	53507	PASS
177	176	5	9	7.1	3779	PASS