

Quantitation Report (QT Reviewed)


Data Path : Z:\VOASRV\HPCHEM1\MSVOA_W\DATA\VW041519\
 Data File : VW009985.D
 Acq On : 15 Apr 2019 21:51
 Operator : SY/VA
 Sample : K2403-15
 Misc : 0.18G/10ML/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 A5128

Manual Integrations
 APPROVED

MMDadoda
 4/17/2019 9:26:25 AM

Quant Time: Apr 16 02:21:27 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM041519S.M
 Quant Title : VOC Analysis
 QLast Update : Tue Apr 16 01:53:26 2019
 Response via : Initial Calibration



Quantitation Report (Qedit)




Data Path : Z:\VOASRV\HPCHEM1\MSVOA_W\DATA\VW041519\
Data File : VW009985.D
Acq On : 15 Apr 2019 21:51
Operator : SY/VA
Sample : K2403-15
Misc : 0.18G/10ML/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampled :
A5128

Quant Time: Apr 16 02:20:30 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM041519S.M
Quant Title : VOC Analysis
QLast Update : Tue Apr 16 01:53:26 2019
Response via : Initial Calibration

Manual Integrations
APPROVED

MMDadoda
4/17/2019 9:26:25 AM



(26) 1,2-Dichloroethane-d4 (S)

8.390min (+0.085) 0.02ug/L

response 236

Ion	Exp%	Act%
65.00	100	100
67.00	50.40	38.98
51.00	99.90	103.81
0.00	0.00	0.00

Quantitation Report (Qedit)




Data Path : Z:\VOASRV\HPCHEM1\MSVOA_W\DATA\VW041519\
 Data File : VW009985.D
 Acq On : 15 Apr 2019 21:51
 Operator : SY/VA
 Sample : K2403-15
 Misc : 0.18G/10ML/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampled :
 A5128

Manual Integrations
 APPROVED

MMDadoda
 4/17/2019 9:26:25 AM

Quant Time: Apr 16 02:20:30 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM041519S.M
 Quant Title : VOC Analysis
 QLast Update : Tue Apr 16 01:53:26 2019
 Response via : Initial Calibration



TIC: VW009985.D

(26) 1,2-Dichloroethane-d4 (S)

8.305min (-0.000) 26.77ug/L m

response 291615

Ion	Exp%	Act%
65.00	100	100
67.00	50.40	0.03#
51.00	99.90	0.08#
0.00	0.00	0.00

Quantitation Report (QT Reviewed)

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_W\DATA\VW041519\
 Data File : VW009985.D
 Acq On : 15 Apr 2019 21:51
 Operator : SY/VA
 Sample : K2403-15
 Misc : 0.18G/10ML/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 Client Sampled :
 A5128

Manual Integrations
 APPROVED

MMDadoda
 4/17/2019 9:26:25 AM

Quant Time: Apr 16 02:21:27 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM041519S.M
 Quant Title : VOC Analysis
 QLast Update : Tue Apr 16 01:53:26 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev (Min)
1) 1,4-Difluorobenzene	8.84	114	607253	25.00	ug/L	0.00
28) Chlorobenzene-d5	11.63	117	605092	25.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.56	152	260640	25.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.36	65	171248	29.84	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	119.36%
7) Chloroethane-d5	2.89	69	146319	25.17	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	100.68%
10) 1,1-Dichloroethene-d2	4.01	63	314909	19.79	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	79.16%
20) 2-Butanone-d5	7.07	46	208130	67.05	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	134.10%
24) Chloroform-d	7.64	84	414946	23.64	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	94.56%
26) 1,2-Dichloroethane-d4	8.31	65	291615m	26.77	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	107.08%
29) Benzene-d6	8.27	84	790015	23.76	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	95.04%
33) 1,2-Dichloropropane-d6	9.27	67	255686	24.14	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	96.56%
37) Toluene-d8	10.32	98	707396	23.32	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	93.28%
38) trans-1,3-Dichloropropene-	10.58	79	126108	25.11	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	100.44%
39) 2-Hexanone-d5	10.93	63	175420	74.74	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	149.48%#
48) 1,1,2,2-Tetrachloroethane-	12.69	84	283656	28.63	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	114.52%
61) 1,2-Dichlorobenzene-d4	13.86	152	260552	26.59	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	106.36%

Target Compounds

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
13) Acetone	4.12	43	6732	2.439	ug/L	60

(#) = qualifier out of range (m) = manual integration (+) = signals summed