



Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW041521\
 Data File : VW018652.D
 Acq On : 15 Apr 2021 11:50
 Operator : SY/VA
 Sample : BFB06
 Misc : 5.00G/10ML/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 BFB106

Integration File: LSCINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM041421SMA.M
 Title : SFAM01.0
 Last Update : Tue Apr 20 01:14:59 2021



AutoFind: Scans 1756, 1757, 1758; Background Corrected with Scan 1749

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.2	12098	PASS
75	95	30	80	54.1	34088	PASS
95	95	100	100	100.0	62973	PASS
96	95	5	9	7.1	4455	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	120	67.0	42219	PASS
175	174	5	9	8.6	3626	PASS
176	174	95	105	97.0	40933	PASS
177	176	5	10	7.5	3077	PASS

m/z	Abundance
38.90	70.0
39.80	235.0
40.90	102.0
44.00	406.0
69.10	112.0
168.80	82.0

Instrument :
MSVOA_W
ClientSampleId :
BFB106

m/z	Abundance
35.90	516.0
37.00	2874.0
38.00	2942.0
39.00	1154.0
40.20	206.0
43.00	121.0
44.00	639.0
45.00	843.0
47.00	750.0
47.90	424.0
49.00	2546.0
50.00	10495.0
51.10	3127.0
52.20	139.0
55.00	227.0
56.00	994.0
57.00	1945.0
58.10	75.0
59.00	349.0
59.90	588.0
61.00	2753.0
62.00	2924.0
63.00	2328.0
64.00	256.0
64.90	229.0
66.10	64.0
67.00	200.0
68.00	6141.0
69.00	5670.0
70.10	513.0
70.90	70.0
72.00	310.0
73.00	5134.0
74.00	10491.0
75.00	29696.0
76.00	2433.0
77.00	376.0
78.00	288.0
78.90	1649.0
79.90	576.0
80.90	2098.0
81.80	495.0
83.00	72.0
85.20	94.0
86.90	2210.0
87.90	2568.0
91.00	318.0
92.00	1634.0
93.00	2490.0
94.00	6587.0
95.00	53880.0
96.00	3817.0
96.90	222.0
102.90	456.0
104.10	202.0
104.90	223.0
105.90	204.0
111.00	248.0
111.70	234.0
114.70	233.0
116.00	258.0
116.80	527.0
118.00	371.0
118.90	750.0
121.00	78.0
122.00	75.0
125.00	769.0
126.10	124.0
127.80	205.0
129.00	107.0
130.00	157.0
130.80	113.0
132.00	117.0
133.00	2129.0
134.10	283.0
134.90	320.0
136.10	53.0
140.90	614.0
141.90	56.0
142.90	556.0
144.70	102.0
145.00	92.0
145.90	72.0
147.00	223.0
148.00	203.0
148.90	116.0
154.70	56.0
155.00	101.0
156.80	97.0
158.30	66.0
159.00	143.0
160.80	169.0

Instrument :
MSVOA_W
ClientSampleId :
BFB106

162.90	292.0
164.10	79.0
164.80	133.0
170.90	89.0
172.10	247.0
174.00	28760.0
174.90	3082.0
175.90	29224.0
176.90	2631.0
177.90	219.0
178.90	674.0
179.90	83.0
189.00	253.0
190.00	99.0
191.00	1820.0
192.00	566.0
193.00	2457.0
194.10	445.0
194.90	350.0
195.80	64.0
203.00	288.0
203.90	95.0
204.90	422.0
207.00	1379.0
207.90	327.0
208.90	113.0
234.80	156.0
235.80	122.0
249.00	1162.0
249.80	290.0
250.30	156.0
251.00	480.0
252.00	122.0
252.90	55.0
264.90	955.0
266.00	341.0
267.00	406.0
277.10	99.0
279.00	204.0
281.00	16760.0
282.00	4862.0
283.10	2406.0
284.00	552.0
285.00	177.0

Instrument :
MSVOA_W
ClientSampleId :
BFB106

m/z	Abundance
36.00	640.0
37.00	3318.0
38.00	3384.0
39.00	1267.0
39.90	140.0
42.90	152.0
44.10	834.0
45.00	810.0
46.10	163.0
46.90	759.0
48.00	561.0
49.00	2823.0
50.00	13705.0
51.00	4155.0
52.00	233.0
55.00	257.0
56.00	1364.0
57.00	2388.0
57.90	144.0
59.00	290.0
60.00	828.0
61.00	3904.0
62.00	3892.0
63.00	3015.0
63.90	290.0
65.00	252.0
66.10	146.0
67.00	142.0
68.00	7795.0
69.00	7584.0
70.00	767.0
71.90	258.0
73.00	4478.0
74.00	12855.0
75.00	37592.0
76.00	3138.0
77.10	596.0
78.00	287.0
78.90	2819.0
79.90	739.0
81.00	2593.0
81.80	478.0
82.90	85.0
85.30	66.0
86.90	3002.0
88.00	2505.0
90.70	182.0
91.90	2176.0
92.90	3158.0
94.00	8115.0
95.00	68928.0
96.00	4876.0
97.00	300.0
103.00	331.0
103.90	591.0
105.00	156.0
105.90	307.0
106.80	90.0
110.90	174.0
111.80	79.0
114.80	144.0
115.80	232.0
116.90	554.0
117.70	483.0
118.90	686.0
123.10	101.0
124.90	505.0
126.00	126.0
127.90	274.0
128.90	86.0
129.90	205.0
131.00	82.0
132.10	91.0
133.00	1587.0
134.00	219.0
135.00	191.0
136.10	80.0
136.90	81.0
140.90	627.0
141.70	152.0
142.90	563.0
144.80	76.0
145.80	97.0
146.80	292.0
148.00	172.0
149.00	82.0
150.00	80.0
154.80	136.0
156.90	143.0
159.10	96.0
160.90	120.0
162.90	210.0

Instrument :
MSVOA_W
ClientSampleId :
BFB106

164.90	193.0
168.90	105.0
170.80	67.0
172.10	379.0
173.90	45032.0
174.90	3903.0
175.90	46680.0
176.90	3674.0
177.90	270.0
179.00	399.0
189.00	108.0
190.10	87.0
190.90	1238.0
192.10	303.0
193.00	1501.0
193.90	376.0
195.00	206.0
203.00	172.0
204.90	243.0
206.00	142.0
207.00	753.0
207.90	195.0
208.90	165.0
218.80	55.0
220.80	69.0
234.80	199.0
248.80	539.0
250.00	228.0
250.90	423.0
252.00	96.0
265.00	1048.0
266.00	299.0
267.00	352.0
277.20	80.0
278.80	261.0
279.40	132.0
281.00	9516.0
282.00	3118.0
283.10	2221.0
284.10	322.0

Instrument :
MSVOA_W
ClientSampleId :
BFB106

m/z	Abundance
36.00	748.0
37.00	2995.0
38.00	3087.0
39.00	1042.0
40.00	304.0
41.00	58.0
42.90	127.0
43.20	137.0
44.00	870.0
45.00	727.0
46.00	96.0
47.00	723.0
48.10	388.0
49.00	2654.0
50.00	12095.0
51.00	3758.0
51.90	268.0
54.90	272.0
56.00	1048.0
57.00	1872.0
57.90	97.0
58.90	122.0
59.90	654.0
61.00	3216.0
62.00	3259.0
63.00	2441.0
64.10	347.0
65.00	343.0
66.10	56.0
67.10	245.0
68.00	7528.0
69.00	6931.0
70.10	577.0
71.90	386.0
73.00	3808.0
74.00	11468.0
75.00	34976.0
76.00	2927.0
77.00	465.0
77.90	408.0
78.90	2391.0
80.00	600.0
81.00	2397.0
81.90	527.0
82.80	86.0
86.90	2369.0
87.90	2568.0
88.90	77.0
90.90	249.0
92.00	2303.0
93.00	2968.0
94.00	7268.0
95.00	66112.0
96.10	4672.0
96.90	109.0
102.90	265.0
104.00	290.0
105.00	118.0
105.90	302.0
106.80	123.0
109.60	98.0
110.00	72.0
111.10	107.0
112.00	117.0
112.90	69.0
114.90	221.0
115.90	223.0
116.90	414.0
117.90	408.0
118.90	583.0
123.80	75.0
125.10	214.0
126.00	51.0
127.90	243.0
128.90	111.0
129.90	181.0
130.90	120.0
133.10	669.0
134.00	109.0
134.70	254.0
135.90	95.0
136.80	95.0
140.90	732.0
141.90	106.0
142.80	646.0
144.70	81.0
145.90	99.0
146.70	122.0
147.90	151.0
149.00	65.0
150.00	87.0
154.90	153.0

Instrument :
MSVOA_W
ClientSampleId :
BFB106

156.80	122.0
158.90	100.0
160.80	105.0
162.90	206.0
164.90	117.0
168.90	103.0
171.90	539.0
173.90	52864.0
174.90	3893.0
175.90	46896.0
176.90	2925.0
177.90	72.0
178.80	267.0
190.90	565.0
191.90	185.0
193.00	732.0
193.90	179.0
195.00	138.0
203.00	141.0
205.00	159.0
206.00	52.0
207.00	626.0
208.20	187.0
208.90	91.0
235.00	95.0
249.00	352.0
249.70	132.0
250.90	297.0
265.00	695.0
266.00	274.0
267.00	166.0
279.00	141.0
279.90	222.0
281.00	5493.0
282.10	2000.0
283.10	1387.0
284.00	202.0
285.10	126.0

Instrument :
MSVOA_W
ClientSampleId :
BFB106