


Data Path : Z:\voasrv\HPCHEM1\MSVOA W\Data\VW041719\
Data File : VW010010.D
Acq On : 17 Apr 2019 09:51
Operator : SY/VA
Sample : BFB
Misc : 5.00G/5ML/MSVOA W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W041719S.M
Title : SW846 8260
Last Update : Thu Apr 18 11:26:35 2019



AutoFind: Scans 1757, 1758, 1759; Background Corrected with Scan 1748

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	23.3	73514	PASS
75	95	30	60	55.0	173525	PASS
95	95	100	100	100.0	315264	PASS
96	95	5	9	6.6	20650	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	76.5	241024	PASS
175	174	5	9	7.9	19112	PASS
176	174	95	101	95.9	231146	PASS
177	176	5	9	6.7	15455	PASS