


Data Path : Z:\voasrv\HPCHEM1\MSVOA W\Data\VW041719\
 Data File : VW010029.D
 Acq On : 17 Apr 2019 21:41
 Operator : SY/VA
 Sample : BFB
 Misc : 5.00G/5ML/MSVOA W/SOIL
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W041719S.M
 Title : SW846 8260
 Last Update : Thu Apr 18 11:26:35 2019



AutoFind: Scans 1757, 1758, 1759; Background Corrected with Scan 1748

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	22.9	66669	PASS
75	95	30	60	55.2	160832	PASS
95	95	100	100	100.0	291242	PASS
96	95	5	9	6.6	19209	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	76.0	221461	PASS
175	174	5	9	8.5	18896	PASS
176	174	95	101	97.3	215509	PASS
177	176	5	9	6.5	13931	PASS