

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW041819\
 Data File : VW010045.D
 Acq On : 18 Apr 2019 15:12
 Operator : SY/VA
 Sample : K2411-10
 Misc : 6.20G/5ML/MSVOA W/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 TP-14-D2

Quant Time: Apr 18 15:57:04 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W041719S.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 18 11:26:35 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	278382	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.84	114	511547	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	454690	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.56	152	197542	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.30	65	219370	63.86	ug/l	0.00
Spiked Amount 50.000			Recovery	=	127.72%	
35) Dibromofluoromethane	7.88	113	173781	53.43	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.86%	
50) Toluene-d8	10.32	98	580175	47.61	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.22%	
62) 4-Bromofluorobenzene	12.62	95	194812	43.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.96%	
Target Compounds						
8) Diethyl Ether	3.67	74	4267	3.007	ug/l	81
16) Acetone	4.12	43	59080	74.120	ug/l	97
20) Methylene Chloride	4.91	84	55687	14.950	ug/l	98
52) Toluene	10.39	92	14553	1.635	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW041819\
Data File : VW010045.D
Acq On : 18 Apr 2019 15:12
Operator : SY/VA
Sample : K2411-10
Misc : 6.20G/5ML/MSVOA W/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
TP-14-D2

Quant Time: Apr 18 15:57:04 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W041719S.M
Quant Title : SW846 8260
QLast Update : Thu Apr 18 11:26:35 2019
Response via : Initial Calibration

