

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW042020\
 Data File : VW015228.D
 Acq On : 20 Apr 2020 17:35
 Operator : SY/VA
 Sample : L2330-01
 Misc : 4.64G/5ML/MSVOA W/SOIL
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 WB-9-1

Quant Time: Apr 21 06:55:05 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W042020S.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 21 06:39:58 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	187735	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.84	114	257913	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	234204	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.56	152	105632	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.31	65	72703	47.36	ug/l	0.00
Spiked Amount			Recovery	=	94.72%	
35) Dibromofluoromethane	7.88	113	74021	51.81	ug/l	0.00
Spiked Amount			Recovery	=	103.62%	
50) Toluene-d8	10.32	98	312430	51.99	ug/l	0.00
Spiked Amount			Recovery	=	103.98%	
62) 4-Bromofluorobenzene	12.62	95	106415	46.41	ug/l	0.00
Spiked Amount			Recovery	=	92.82%	


Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
11) Tert butyl alcohol	5.17	59	13951	111.480	ug/l #	94
13) Acrolein	3.92	56	375	5.725	ug/l #	34
16) Acetone	4.13	43	31458	117.319	ug/l	95
17) Carbon Disulfide	4.39	76	9962	2.093	ug/l	98
19) Methyl tert-butyl Ether	5.43	73	5731	1.559	ug/l #	86
20) Methylene Chloride	4.93	84	4887	2.781	ug/l	89
25) 2-Butanone	7.18	43	16532	46.471	ug/l	95
29) Tetrahydrofuran	7.54	42	286	1.290	ug/l #	62
31) Cyclohexane	7.95	56	4153	1.463	ug/l #	16
36) 1,1-Dichloropropene	7.95	75	12409	5.331	ug/l #	50
37) Ethyl Acetate	7.18	43	16532	21.883	ug/l #	67
38) Carbon Tetrachloride	7.95	117	17018	6.690	ug/l #	19
40) Benzene	8.33	78	10354	1.665	ug/l	99
48) Methyl methacrylate	9.40	41	1130	1.509	ug/l #	42
51) 4-Methyl-2-Pentanone	10.21	43	1562	2.100	ug/l	100
55) 1,1,2-Trichloroethane	10.76	97	2178	1.859	ug/l #	24
59) 2-Hexanone	10.83	43	50291	94.718	ug/l	75
73) Isopropylbenzene	12.46	105	291193	39.759	ug/l	99
74) N-amyl acetate	12.25	43	5723	4.492	ug/l #	50
76) 1,2,3-Trichloropropane	12.62	75	53406	65.354	ug/l #	100
81) trans-1,4-Dichloro-2-buten	12.46	75	3201	8.563	ug/l #	35
84) 1,2,4-Trimethylbenzene	13.25	105	34266	5.507	ug/l	98
85) sec-Butylbenzene	13.38	105	11052	1.451	ug/l	84
86) p-Isopropyltoluene	13.50	119	9231	1.305	ug/l	95
90) Hexachloroethane	14.09	117	2790	2.150	ug/l #	16
95) Naphthalene	15.36	128	23672	7.287	ug/l	97
96) 1,2,3-Trichlorobenzene	15.56	180	1963	1.195	ug/l #	59


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW042020\
Data File : VW015228.D
Acq On : 20 Apr 2020 17:35
Operator : SY/VA
Sample : L2330-01
Misc : 4.64G/5ML/MSVOA W/SOIL
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
WB-9-1

Quant Time: Apr 21 06:55:05 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W042020S.M
Quant Title : SW846 8260
QLast Update : Tue Apr 21 06:39:58 2020
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.95 min Scan# 992

Delta R.T. 0.00 min

Lab File: VW015228.D

Acq: 20 Apr 2020 17:35

Instrument :

MSVOA_W

ClientSampleId :


WB-9-1

Tot Ion:168 Resp: 187735

Ion Ratio Lower Upper

168 100

99 52.8 48.5 72.7



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VW

7.95

80000



60000

40000

20000

0

Time--> 7.80 7.90 8.00 8.10



#11

Tert butyl alcohol

Concen: 111.480 ug/l

RT: 5.17 min Scan# 535

Delta R.T. -0.01 min

Lab File: VW015228.D


Acq: 20 Apr 2020 17:35

Tgt Ion: 59 Resp: 13951

Ion Ratio Lower Upper

59 100

57 12.9 8.4 12.6#



Abundance Ion 59.00 (58.70 to 59.70): VW

Ion 57.00 (56.70 to 57.70): VW

5.17

4000


3000


2000

1000

0

Time--> 5.00 5.10 5.20 5.30





#13

Acrolein

Concen: 5.725 ug/l

RT: 3.92 min Scan# 330

Delta R.T. 0.01 min

Lab File: VW015228.D

Acq: 20 Apr 2020 17:35

Instrument :

MSVOA_W

ClientSampled :





WB-9-1

Tot Ion: 56 Resp: 375

Ion Ratio Lower Upper

56 100

55 16.3 56.2 84.2#



#16

Acetone

Concen: 117.319 ug/l

RT: 4.13 min Scan# 365

Delta R.T. 0.00 min

Lab File: VW015228.D




Acq: 20 Apr 2020 17:35


Tgt Ion: 43 Resp: 31458

Ion Ratio Lower Upper

43 100

58 30.7 27.0 40.4









#17
Carbon Disulfide
Concen: 2.093 ug/l
RT: 4.39 min Scan# 408
Delta R.T. 0.01 min
Lab File: VW015228.D
Acq: 20 Apr 2020 17:35




Instrument :
MSVOA_W
ClientSampled :
WB-9-1


Tot Ion: 76 Resp: 9962
Ion Ratio Lower Upper
76 100
78 9.6 7.0 10.6



#19
Methyl tert-butyl Ether
Concen: 1.559 ug/l
RT: 5.43 min Scan# 579
Delta R.T. 0.01 min
Lab File: VW015228.D
Acq: 20 Apr 2020 17:35

Tgt Ion: 73 Resp: 5731
Ion Ratio Lower Upper
73 100
57 28.7 17.8 26.6#









#20
Methylene Chloride
Concen: 2.781 ug/l
RT: 4.93 min Scan# 496
Delta R.T. 0.01 min
Lab File: VW015228.D
Acq: 20 Apr 2020 17:35




Instrument :
MSVOA_W
ClientSampled :
WB-9-1


Tot Ion:	84	Resp:	4887
Ion	Ratio	Lower	Upper
84	100		
49	98.4	95.4	143.0
51	35.1	29.8	44.8
86	60.2	47.4	71.0



#25
2-Butanone
Concen: 46.471 ug/l
RT: 7.18 min Scan# 865
Delta R.T. 0.01 min
Lab File: VW015228.D
Acq: 20 Apr 2020 17:35

Tgt Ion:	43	Resp:	16532
Ion	Ratio	Lower	Upper
43	100		
72	29.6	21.6	32.4









#29
Tetrahydrofuran
Concen: 1.290 ug/l
RT: 7.54 min Scan# 925
Delta R.T. 0.01 min
Lab File: VW015228.D
Acq: 20 Apr 2020 17:35




Instrument :
MSVOA_W
ClientSampleId :
WB-9-1


Tot Ion: 42 Resp: 286
Ion Ratio Lower Upper
42 100
72 0.0 36.2 54.2#
71 46.2 34.2 51.4



#31
Cyclohexane
Concen: 1.463 ug/l
RT: 7.95 min Scan# 992
Delta R.T. -0.01 min
Lab File: VW015228.D
Acq: 20 Apr 2020 17:35

Tgt Ion: 56 Resp: 4153
Ion Ratio Lower Upper
56 100
69 156.7 27.1 40.7#
84 122.1 71.9 107.9#









#33
 1,2-Dichloroethane-d4
 Concen: 47.357 ug/l
 RT: 8.31 min Scan# 1051
 Delta R.T. 0.01 min
 Lab File: VW015228.D
 Acq: 20 Apr 2020 17:35




Instrument :
 MSVOA_W
 ClientSampleId :
 WB-9-1


Tot Ion: 65 Resp: 72703
 Ion Ratio Lower Upper
 65 100
 67 52.1 0.0 107.0



#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 8.84 min Scan# 1138
 Delta R.T. 0.00 min
 Lab File: VW015228.D
 Acq: 20 Apr 2020 17:35

Tgt Ion: 114 Resp: 257913
 Ion Ratio Lower Upper
 114 100
 63 18.5 0.0 39.8
 88 15.1 0.0 30.2





#35

Dibromofluoromethane

Concen: 51.813 ug/l

RT: 7.88 min Scan# 981

Delta R.T. 0.00 min

Lab File: VW015228.D

Acq: 20 Apr 2020 17:35

Instrument :

MSVOA_W

ClientSampleId :

WB-9-1


Tot Ion:113 Resp: 74021

Ion Ratio Lower Upper

113 100

111 100.9 83.0 124.4




192 23.0 15.1 22.7#



Abundance Ion 112.80 (112.50 to 113.50): V

Ion 110.80 (110.50 to 111.50): V

Ion 191.70 (191.40 to 192.40): V



#36

1,1-Dichloropropene

Concen: 5.331 ug/l

RT: 7.95 min Scan# 992

Delta R.T. -0.13 min

Lab File: VW015228.D

Acq: 20 Apr 2020 17:35


Tgt Ion: 75 Resp: 12409

Ion Ratio Lower Upper

75 100

110 9.4 18.2 54.6#



77 0.0 24.5 36.7#




Abundance Ion 74.90 (74.60 to 75.60): VW

Ion 109.90 (109.60 to 110.60): V

Ion 76.90 (76.60 to 77.60): VW









#37
Ethyl Acetate
Concen: 21.883 ug/l
RT: 7.18 min Scan# 865
Delta R.T. -0.08 min
Lab File: VW015228.D
Acq: 20 Apr 2020 17:35




Instrument :
MSVOA_W
ClientSampleId :
WB-9-1


Tot Ion: 43 Resp: 16532
Ion Ratio Lower Upper
43 100
61 0.0 11.9 17.9#
70 0.1 8.8 13.2#



#38
Carbon Tetrachloride
Concen: 6.690 ug/l
RT: 7.95 min Scan# 992
Delta R.T. -0.12 min
Lab File: VW015228.D
Acq: 20 Apr 2020 17:35

Tgt Ion: 117 Resp: 17018
Ion Ratio Lower Upper
117 100
119 5.2 72.3 108.5#
121 0.0 23.7 35.5#





#40

Benzene

Concen: 1.665 ug/l

RT: 8.33 min Scan# 1054

Delta R.T. 0.01 min

Lab File: VW015228.D

Acq: 20 Apr 2020 17:35

Instrument :

MSVOA_W

ClientSampled :





WB-9-1

Tot Ion: 78 Resp: 10354

Ion Ratio Lower Upper

78 100

77 23.0 18.2 27.2



#48

Methyl methacrylate

Concen: 1.509 ug/l

RT: 9.40 min Scan# 1229

Delta R.T. -0.04 min

Lab File: VW015228.D

Acq: 20 Apr 2020 17:35




Tgt Ion: 41 Resp: 1130


Ion Ratio Lower Upper

41 100

69 8.8 70.6 105.8#

39 44.8 42.6 64.0









#50
Toluene-d8
Concen: 51.987 ug/l
RT: 10.32 min Scan# 1381
Delta R.T. 0.00 min
Lab File: VW015228.D
Acq: 20 Apr 2020 17:35




Instrument :
MSVOA_W
ClientSampled :
WB-9-1


Tot Ion: 98 Resp: 312430
Ion Ratio Lower Upper
98 100
100 67.0 52.9 79.3



#51
4-Methyl-2-Pentanone
Concen: 2.100 ug/l
RT: 10.21 min Scan# 1362
Delta R.T. 0.00 min
Lab File: VW015228.D
Acq: 20 Apr 2020 17:35

Tgt Ion: 43 Resp: 1562
Ion Ratio Lower Upper
43 100
58 40.3 32.3 48.5









#55
 1,1,2-Trichloroethane
 Concen: 1.859 ug/l
 RT: 10.76 min Scan# 1452
 Delta R.T. -0.03 min
 Lab File: VW015228.D
 Acq: 20 Apr 2020 17:35




Instrument :
 MSVOA_W
 ClientSampled :
 WB-9-1


Tot Ion: 97 Resp: 2178
 Ion Ratio Lower Upper
 97 100
 83 7.8 67.3 100.9#
 85 13.1 43.8 65.8#
 99 0.0 51.4 77.0#



#59
 2-Hexanone
 Concen: 94.718 ug/l
 RT: 10.83 min Scan# 1464
 Delta R.T. -0.14 min
 Lab File: VW015228.D
 Acq: 20 Apr 2020 17:35

Tgt Ion: 43 Resp: 50291
 Ion Ratio Lower Upper
 43 100
 58 38.2 28.1 84.3









#62
4-Bromofluorobenzene
Concen: 46.408 ug/l
RT: 12.62 min Scan# 1757
Delta R.T. 0.00 min
Lab File: VW015228.D
Acq: 20 Apr 2020 17:35




Instrument :
MSVOA_W
ClientSampled :
WB-9-1


Tot Ion	95	Resp	106415
Ion Ratio	100		
Lower		0.0	144.4
Upper		0.0	140.6



#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.63 min Scan# 1596
Delta R.T. 0.00 min
Lab File: VW015228.D
Acq: 20 Apr 2020 17:35

Tgt Ion	117	Resp	234204
Ion Ratio	100		
Lower		44.1	66.1
Upper		24.9	37.3





#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.56 min Scan# 1912

Delta R.T. 0.00 min

Lab File: VW015228.D

Acq: 20 Apr 2020 17:35

Instrument :

MSVOA_W

ClientSampleId :

WB-9-1


Tot Ion:152 Resp: 105632

Ion Ratio Lower Upper

152 100

115 56.1 31.3 93.9




150 156.1 0.0 343.4



Abundance Ion 151.90 (151.60 to 152.60): V

Ion 114.90 (114.60 to 115.60): V

Ion 149.90 (149.60 to 150.60): V



#73

Isopropylbenzene

Concen: 39.759 ug/l

RT: 12.46 min Scan# 1732

Delta R.T. 0.00 min

Lab File: VW015228.D


Acq: 20 Apr 2020 17:35

Tgt Ion:105 Resp: 291193

Ion Ratio Lower Upper



105 100


120 27.1 13.7 41.1



Abundance Ion 105.00 (104.70 to 105.70): V

Ion 120.00 (119.70 to 120.70): V









#74
N-amvl acetate
Concen: 4.492 ug/l
RT: 12.25 min Scan# 1697
Delta R.T. -0.02 min
Lab File: VW015228.D
Acq: 20 Apr 2020 17:35

Instrument :
MSVOA_W
ClientSampled :
WB-9-1

Tot Ion: 43 Resp: 5723
Ion Ratio Lower Upper
43 100
70 11.7 35.4 53.2#
55 0.0 18.5 27.7#
61 0.0 20.3 30.5#




Abundance Ion 43.00 (42.70 to 43.70): VW
Ion 70.00 (69.70 to 70.70): VW
Ion 55.00 (54.70 to 55.70): VW
Ion 61.00 (60.70 to 61.70): VW






#76
1,2,3-Trichloropropane
Concen: 65.354 ug/l
RT: 12.62 min Scan# 1757
Delta R.T. -0.15 min
Lab File: VW015228.D
Acq: 20 Apr 2020 17:35

Tgt Ion: 75 Resp: 53406
Ion Ratio Lower Upper
75 100
77 1.5 0.0 0.0#



Abundance Ion 74.90 (74.60 to 75.60): VW
Ion 77.00 (76.70 to 77.70): VW





#81

trans-1,4-Dichloro-2-butene

Concen: 8.563 ug/l

RT: 12.46 min Scan# 1732

Delta R.T. -0.05 min

Lab File: VW015228.D

Acq: 20 Apr 2020 17:35

Instrument :

MSVOA_W

ClientSampled :

WB-9-1


Tot Ion: 75 Resp: 3201

Ion Ratio Lower Upper




75 100

53 124.7 69.4 104.2#

89 115.4 34.8 52.2#



Abundance Ion 74.90 (74.60 to 75.60): VW
Ion 53.00 (52.70 to 53.70): VW
Ion 88.90 (88.60 to 89.60): VW



#84

1,2,4-Trimethylbenzene

Concen: 5.507 ug/l

RT: 13.25 min Scan# 1861

Delta R.T. 0.00 min

Lab File: VW015228.D


Acq: 20 Apr 2020 17:35

Tgt Ion: 105 Resp: 34266



Ion Ratio Lower Upper


105 100

120 44.8 23.0 69.0



Abundance Ion 105.00 (104.70 to 105.70): V
Ion 120.00 (119.70 to 120.70): V





#85

sec-Butylbenzene

Concen: 1.451 ug/l

RT: 13.38 min Scan# 1883

Delta R.T. 0.00 min

Lab File: VW015228.D

Acq: 20 Apr 2020 17:35

Instrument :

MSVOA_W

ClientSampled :


WB-9-1

Tot Ion:105 Resp: 11052

Ion Ratio Lower Upper




105 100

134 27.6 10.1 30.1



Abundance Ion 105.00 (104.70 to 105.70): V

Ion 134.00 (133.70 to 134.70): V



#86

p-Isopropyltoluene

Concen: 1.305 ug/l

RT: 13.50 min Scan# 1902

Delta R.T. 0.00 min

Lab File: VW015228.D

Acq: 20 Apr 2020 17:35


Tgt Ion:119 Resp: 9231

Ion Ratio Lower Upper

119 100

134 26.6 13.3 39.9



91 29.8 12.0 36.1




Abundance Ion 119.00 (118.70 to 119.70): V

Ion 134.00 (133.70 to 134.70): V

Ion 91.00 (90.70 to 91.70): V





#90

Hexachloroethane

Concen: 2.150 ug/l

RT: 14.09 min Scan# 1999

Delta R.T. 0.00 min

Lab File: VW015228.D

Acq: 20 Apr 2020 17:35

Instrument :

MSVOA_W

ClientSampled :


WB-9-1

Tot Ion:117 Resp: 2790

Ion Ratio Lower Upper




117 100

201 0.0 33.6 100.6#



Abundance Ion 116.80 (116.50 to 117.50): V

Ion 200.70 (200.40 to 201.40): V



#95

Naphthalene

Concen: 7.287 ug/l

RT: 15.36 min Scan# 2208

Delta R.T. 0.00 min

Lab File: VW015228.D

Acq: 20 Apr 2020 17:35


Tgt Ion:128 Resp: 23672

Ion Ratio Lower Upper

128 100

127 11.5 10.3 15.5



129 12.0 8.6 12.8




Abundance Ion 128.00 (127.70 to 128.70): V

Ion 127.00 (126.70 to 127.70): V

Ion 129.00 (128.70 to 129.70): V






#96
 1,2,3-Trichlorobenzene
 Concen: 1.195 ug/l
 RT: 15.56 min Scan# 2240
 Delta R.T. 0.01 min
 Lab File: VW015228.D
 Acq: 20 Apr 2020 17:35

Instrument :
 MSVOA_W
 ClientSampleId :
 WB-9-1

Tot Ion:180 Resp: 1963
 Ion Ratio Lower Upper
 180 100
 182 63.5 48.8 146.3
 145 0.0 17.2 51.6#



Abundance Ion 179.80 (179.50 to 180.50): V
 Ion 181.80 (181.50 to 182.50): V
 Ion 144.90 (144.60 to 145.60): V

