

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW042020\
 Data File : VW015243.D
 Acq On : 20 Apr 2020 23:54
 Operator : SY/VA
 Sample : L2304-07
 Misc : 5.04G/5ML/MSVOA W/SOIL
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 WC-B2-VOA

Quant Time: Apr 21 08:05:24 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W042020S.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 21 06:39:58 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	199282	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.84	114	283224	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	256627	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.56	152	115332	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	80315	49.28	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.56%	
35) Dibromofluoromethane	7.88	113	79965	50.97	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.94%	
50) Toluene-d8	10.32	98	341022	51.67	ug/l	0.00
Spiked Amount	50.000		Recovery	=	103.34%	
62) 4-Bromofluorobenzene	12.62	95	116442	46.24	ug/l	0.00
Spiked Amount	50.000		Recovery	=	92.48%	
Target Compounds						
16) Acetone	4.14	43	3062	10.758	ug/l #	87
20) Methylene Chloride	4.92	84	14466	7.755	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW042020\
Data File : VW015243.D
Acq On : 20 Apr 2020 23:54
Operator : SY/VA
Sample : L2304-07
Misc : 5.04G/5ML/MSVOA W/SOIL
ALS Vial : 30 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
WC-B2-VOA

Quant Time: Apr 21 08:05:24 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W042020S.M
Quant Title : SW846 8260
QLast Update : Tue Apr 21 06:39:58 2020
Response via : Initial Calibration

