


Data Path : Z:\voasrv\HPCHEM1\MSVOA W\Data\VW042020\
 Data File : VW015223.D
 Acq On : 20 Apr 2020 15:22
 Operator : SY/VA
 Sample : BFB
 Misc : 5.00G/5ML/MSVOA W/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W042020S.M
 Title : SW846 8260
 Last Update : Tue Apr 21 06:39:58 2020



AutoFind: Scans 1756, 1757, 1758; Background Corrected with Scan 1748

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	18.0	8329	PASS
75	95	30	60	51.1	23706	PASS
95	95	100	100	100.0	46397	PASS
96	95	5	9	6.6	3041	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	84.4	39160	PASS
175	174	5	9	7.1	2786	PASS
176	174	95	101	99.4	38936	PASS
177	176	5	9	7.0	2721	PASS