

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW042122\
 Data File : VW022600.D
 Acq On : 21 Apr 2022 12:08
 Operator : SY/VA
 Sample : VW0421SBL01
 Misc : 5.00g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :

Quant Time: Apr 22 01:17:59 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM040722SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Apr 22 01:17:16 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	335916	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.628	117	322194	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	167356	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.355	65	123876	16.681	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	66.720%
7) Chloroethane-d5	2.885	69	77501	16.652	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	66.600%
11) 1,1-Dichloroethene-d2	4.019	63	134063	13.919	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	55.680%
21) 2-Butanone-d5	7.074	46	47107	36.334	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	72.660%
24) Chloroform-d	7.647	84	199838	19.033	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	76.120%
26) 1,2-Dichloroethane-d4	8.378	65	459	0.077	ug/L	0.07
Spiked Amount	25.000	Range	70 - 130	Recovery	=	0.320%#
32) Benzene-d6	8.275	84	393183	19.974	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	79.880%
36) 1,2-Dichloropropane-d6	9.274	67	109176	19.165	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	76.680%
41) Toluene-d8	10.323	98	377635	19.588	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	78.360%
43) trans-1,3-Dichloroprop...	10.573	79	43207	16.071	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	64.280%
47) 2-Hexanone-d5	10.920	63	38180	36.024	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	72.040%
56) 1,1,2,2-Tetrachloroeth...	12.688	84	90375	18.658	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	74.640%
66) 1,2-Dichlorobenzene-d4	13.853	152	119155	20.101	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	80.400%
Target Compounds					Qvalue	
13) Acetone	4.135	43	3803	4.308	ug/L	81


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW042122\
Data File : VW022600.D
Acq On : 21 Apr 2022 12:08
Operator : SY/VA
Sample : VW0421SBL01
Misc : 5.00g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :


Quant Time: Apr 22 01:17:59 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM040722SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Apr 22 01:17:16 2022
Response via : Initial Calibration





#13
 Acetone
 Concen: 4.308 ug/L
 RT: 4.135 min Scan# 30
 Delta R.T. 0.018 min
 Lab File: VW022600.D
 Acq: 21 Apr 2022 12:08

Instrument :
 MSVOA_W
 ClientSampleId :



Tgt Ion: 43 Resp: 3803
 Ion Ratio Lower Upper
 43 100
 58 18.7 0.0 57.0

