

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW042221\
 Data File : VW018730.D
 Acq On : 22 Apr 2021 12:42
 Operator : SY/VA
 Sample : VIBLK113
 Misc : 5.00G/10ML/MSVOA_W/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VIBLK113

Quant Time: Apr 23 02:01:49 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM041421SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Apr 23 01:53:36 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	581628	25.00	ug/L	0.00
28) Chlorobenzene-d5	11.634	117	521560	25.00	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.560	152	252089	25.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.355	65	146333	16.56	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	66.24%
7) Chloroethane-d5	2.891	69	124474	15.89	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	63.56%
11) 1,1-Dichloroethene-d2	4.019	63	228058	14.86	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	59.44%
21) 2-Butanone-d5	7.080	46	95614	47.52	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	95.04%
24) Chloroform-d	7.647	84	359560	22.53	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	90.12%
26) 1,2-Dichloroethane-d4	8.305	65	210664	23.47	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	93.88%
32) Benzene-d6	8.275	84	708051	23.14	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	92.56%
36) 1,2-Dichloropropane-d6	9.274	67	217749	23.62	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	94.48%
41) Toluene-d8	10.323	98	657361	23.49	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	93.96%
43) trans-1,3-Dichloroprop...	10.579	79	93891	23.92	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	95.68%
47) 2-Hexanone-d5	10.927	63	80513	50.22	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	100.44%
56) 1,1,2,2-Tetrachloroeth...	12.694	84	181894	23.62	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	94.48%
66) 1,2-Dichlorobenzene-d4	13.853	152	220900	24.94	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	99.76%
Target Compounds					Qvalue	
42) Toluene	10.390	91	12191	0.35	ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW042221\
Data File : VW018730.D
Acq On : 22 Apr 2021 12:42
Operator : SY/VA
Sample : VIBLK113
Misc : 5.00G/10ML/MSVOA_W/SOIL
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VIBLK113

Quant Time: Apr 23 02:01:49 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM041421SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Apr 23 01:53:36 2021
Response via : Initial Calibration

