

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW042221\
 Data File : VW018741.D
 Acq On : 22 Apr 2021 17:30
 Operator : SY/VA
 Sample : M2121-07
 Misc : 4.21G/10ML/MSVOA_W/SOIL
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 BG505

Quant Time: Apr 23 02:04:02 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM041421SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Apr 23 01:53:36 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	574281	25.00	ug/L	0.00
28) Chlorobenzene-d5	11.634	117	480625	25.00	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.560	152	189079	25.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.355	65	104617	11.99	ug/L	0.00
Spiked Amount	25.000	Range 30 - 150	Recovery =	47.96%		
7) Chloroethane-d5	2.891	69	96410	12.46	ug/L	0.00
Spiked Amount	25.000	Range 30 - 150	Recovery =	49.84%		
11) 1,1-Dichloroethene-d2	4.019	63	135009	8.91	ug/L	0.00
Spiked Amount	25.000	Range 45 - 110	Recovery =	35.64%#		
21) 2-Butanone-d5	7.080	46	92580	46.60	ug/L	0.00
Spiked Amount	50.000	Range 20 - 135	Recovery =	93.20%		
24) Chloroform-d	7.647	84	303522	19.26	ug/L	0.00
Spiked Amount	25.000	Range 40 - 150	Recovery =	77.04%		
26) 1,2-Dichloroethane-d4	8.305	65	171923	19.40	ug/L	0.00
Spiked Amount	25.000	Range 70 - 130	Recovery =	77.60%		
32) Benzene-d6	8.275	84	509705	18.08	ug/L	0.00
Spiked Amount	25.000	Range 20 - 135	Recovery =	72.32%		
36) 1,2-Dichloropropane-d6	9.274	67	180682	21.26	ug/L	0.00
Spiked Amount	25.000	Range 70 - 120	Recovery =	85.04%		
41) Toluene-d8	10.323	98	419749	16.28	ug/L	0.00
Spiked Amount	25.000	Range 30 - 130	Recovery =	65.12%		
43) trans-1,3-Dichloroprop...	10.579	79	55420	15.32	ug/L	0.00
Spiked Amount	25.000	Range 30 - 135	Recovery =	61.28%		
47) 2-Hexanone-d5	10.927	63	75840	51.33	ug/L	0.00
Spiked Amount	50.000	Range 20 - 135	Recovery =	102.66%		
56) 1,1,2,2-Tetrachloroeth...	12.694	84	158550	22.34	ug/L	0.00
Spiked Amount	25.000	Range 45 - 120	Recovery =	89.36%		
66) 1,2-Dichlorobenzene-d4	13.853	152	143789	21.65	ug/L	0.00
Spiked Amount	25.000	Range 75 - 120	Recovery =	86.60%		
Target Compounds						
13) Acetone	4.135	43	7818	5.42	ug/L	99
14) Carbon disulfide	4.385	76	28063	1.25	ug/L	98
16) Methylene chloride	4.928	84	17029	1.81	ug/L	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW042221\
Data File : VW018741.D
Acq On : 22 Apr 2021 17:30
Operator : SY/VA
Sample : M2121-07
Misc : 4.21G/10ML/MSVOA_W/SOIL
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BG505

Quant Time: Apr 23 02:04:02 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM041421SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Apr 23 01:53:36 2021
Response via : Initial Calibration

