

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW042320\  
 Data File : VW015322.D  
 Acq On : 23 Apr 2020 10:41  
 Operator : SY/VA  
 Sample : VW0423SBL01  
 Misc : 5.00G/5ML/MSVOA W/SOIL  
 ALS Vial : 3 Sample Multiplier: 1

Instrument :  
 MSVOA\_W  
 ClientSampleId :  
 VW0423SBL01

Quant Time: Apr 24 07:01:21 2020  
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA\_W\METHOD\82W042020S.M  
 Quant Title : SW846 8260  
 QLast Update : Tue Apr 21 06:39:58 2020  
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	217506	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.84	114	295985	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	262292	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.56	152	114228	50.00	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	8.31	65	85619	48.14	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	96.28%	
35) Dibromofluoromethane	7.88	113	85781	52.32	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	104.64%	
50) Toluene-d8	10.32	98	354865	51.45	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	102.90%	
62) 4-Bromofluorobenzene	12.62	95	116805	44.39	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	88.78%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW042320\  
 Data File : VW015322.D  
 Acq On : 23 Apr 2020 10:41  
 Operator : SY/VA  
 Sample : VW0423SBL01  
 Misc : 5.00G/5ML/MSVOA W/SOIL  
 ALS Vial : 3 Sample Multiplier: 1

Instrument :  
 MSVOA\_W  
 ClientSampleId :  
 VW0423SBL01

Quant Time: Apr 24 07:01:21 2020  
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA\_W\METHOD\82W042020S.M  
 Quant Title : SW846 8260  
 QLast Update : Tue Apr 21 06:39:58 2020  
 Response via : Initial Calibration

