

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW042419\
 Data File : VW010136.D
 Acq On : 25 Apr 2019 11:21
 Operator : SY/VA
 Sample : VW0424SBL03
 Misc : 5.00G/10ML/MSVOA W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VBLK14

Quant Time: Apr 26 01:15:59 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM042419S.M
 Quant Title : VOC Analysis
 QLast Update : Fri Apr 26 01:05:00 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.84	114	805854	25.00	ug/L	0.00
28) Chlorobenzene-d5	11.63	117	752235	25.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.56	152	352739	25.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.35	65	204049	26.49	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	105.96%
7) Chloroethane-d5	2.89	69	161671	25.40	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	101.60%
10) 1,1-Dichloroethene-d2	4.01	63	395386	17.49	ug/L	-0.01
Spiked Amount	25.000	Range	45 - 110	Recovery	=	69.96%
20) 2-Butanone-d5	7.08	46	159608	43.88	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	87.76%
24) Chloroform-d	7.64	84	492366	22.95	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	91.80%
26) 1,2-Dichloroethane-d4	8.30	65	285452	22.75	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	91.00%
29) Benzene-d6	8.27	84	1030967	24.20	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	96.80%
33) 1,2-Dichloropropane-d6	9.27	67	317373	23.89	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	95.56%
37) Toluene-d8	10.32	98	937924	24.24	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	96.96%
38) trans-1,3-Dichloropropene-	10.58	79	138434	22.62	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	90.48%
39) 2-Hexanone-d5	10.93	63	125698	45.04	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	90.08%
48) 1,1,2,2-Tetrachloroethane-	12.69	84	246699	22.38	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	89.52%
61) 1,2-Dichlorobenzene-d4	13.86	152	330116	25.28	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	101.12%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW042419\
Data File : VW010136.D
Acq On : 25 Apr 2019 11:21
Operator : SY/VA
Sample : VW0424SBL03
Misc : 5.00G/10ML/MSVOA W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VBLK14

Quant Time: Apr 26 01:15:59 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM042419S.M
Quant Title : VOC Analysis
QLast Update : Fri Apr 26 01:05:00 2019
Response via : Initial Calibration

