

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW042420\
 Data File : VW015333.D
 Acq On : 24 Apr 2020 10:37
 Operator : SY/VA
 Sample : VSTDCCC050
 Misc : 5.00G/5ML/MSVOA W/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_W
 LabSampleId :
 VSTDCCC050

Quant Time: Apr 25 06:54:37 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W042020S.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 21 06:39:58 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	107	0.00
2 T	Dichlorodifluoromethane	0.395	0.357	9.6	99	0.00
3 P	Chloromethane	0.430	0.375	12.8	96	0.00
4 C	Vinyl Chloride	0.508	0.455	10.4#	98	0.00
5 T	Bromomethane	0.361	0.356	1.4	110	0.00
6 T	Chloroethane	0.315	0.272	13.7	98	0.00
7 T	Trichlorofluoromethane	0.718	0.672	6.4	108	0.00
8 T	Diethyl Ether	0.202	0.178	11.9	101	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.443	0.411	7.2	109	0.00
10 T	Methyl Iodide	0.465	0.427	8.2	93	0.00
11 T	Tert butyl alcohol	0.033	0.028	15.2	110	0.00
12 CM	1,1-Dichloroethene	0.426	0.392	8.0#	107	0.00
13 T	Acrolein	0.017	0.016	5.9	111	0.00
14 T	Allyl chloride	0.566	0.488	13.8	98	0.00
15 T	Acrylonitrile	0.077	0.067	13.0	100	0.00
16 T	Acetone	0.071	0.070	1.4	111	0.00
17 T	Carbon Disulfide	1.267	1.102	13.0	97	0.00
18 T	Methyl Acetate	0.189	0.157	16.9	101	0.00
19 T	Methyl tert-butyl Ether	0.979	0.887	9.4	103	0.00
20 T	Methylene Chloride	0.468	0.403	13.9	100	0.00
21 T	trans-1,2-Dichloroethene	0.484	0.450	7.0	106	0.00
22 T	Diisopropyl ether	1.085	0.988	8.9	103	0.00
23 T	Vinyl Acetate	0.637	0.569	10.7	101	0.00
24 P	1,1-Dichloroethane	0.752	0.701	6.8	106	0.00
25 T	2-Butanone	0.095	0.090	5.3	110	0.00
26 T	2,2-Dichloropropane	0.791	0.730	7.7	109	0.00
27 T	cis-1,2-Dichloroethene	0.522	0.495	5.2	109	0.00
28 T	Bromochloromethane	0.273	0.253	7.3	101	0.00
29 T	Tetrahydrofuran	0.059	0.052	11.9	101	0.00
30 C	Chloroform	0.812	0.771	5.0#	108	0.00
31 T	Cyclohexane	0.756	0.638	15.6	101	0.00
32 T	1,1,1-Trichloroethane	0.784	0.744	5.1	110	0.00
33 S	1,2-Dichloroethane-d4	0.409	0.369	9.8	97	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	106	0.00
35 S	Dibromofluoromethane	0.277	0.271	2.2	102	0.00
36 T	1,1-Dichloropropene	0.451	0.437	3.1	106	0.00
37 T	Ethyl Acetate	0.146	0.132	9.6	101	0.00
38 T	Carbon Tetrachloride	0.493	0.485	1.6	109	0.00
39 T	Methylcyclohexane	0.601	0.550	8.5	102	0.00
40 TM	Benzene	1.205	1.152	4.4	104	0.00
41 T	Methacrylonitrile	0.088	0.087	1.1	121	0.00
42 TM	1,2-Dichloroethane	0.355	0.340	4.2	106	0.00
43 T	Isopropyl Acetate	0.307	0.276	10.1	100	0.00
44 TM	Trichloroethene	0.369	0.374	-1.4	111	0.00
45 C	1,2-Dichloropropane	0.280	0.269	3.9#	106	0.00

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW042420\
 Data File : VW015333.D
 Acq On : 24 Apr 2020 10:37
 Operator : SY/VA
 Sample : VSTDCCC050
 Misc : 5.00G/5ML/MSVOA W/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_W
 LabSampleId :
 VSTDCCC050

Quant Time: Apr 25 06:54:37 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W042020S.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 21 06:39:58 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.162	0.157	3.1	109	0.00
47 T	Bromodichloromethane	0.434	0.419	3.5	107	0.00
48 T	Methyl methacrylate	0.145	0.124	14.5	91	0.00
49 T	1,4-Dioxane	0.002	0.002	0.0	110	0.00
50 S	Toluene-d8	1.165	1.086	6.8	96	0.00
51 T	4-Methyl-2-Pentanone	0.144	0.131	9.0	101	0.00
52 CM	Toluene	0.844	0.820	2.8#	106	0.00
53 T	t-1,3-Dichloropropene	0.427	0.404	5.4	103	0.00
54 T	cis-1,3-Dichloropropene	0.497	0.479	3.6	106	0.00
55 T	1,1,2-Trichloroethane	0.227	0.218	4.0	107	0.00
56 T	Ethyl methacrylate	0.293	0.266	9.2	100	0.00
57 T	1,3-Dichloropropane	0.376	0.362	3.7	106	0.00
58 T	2-Chloroethyl Vinyl ether	0.136	0.132	2.9	102	0.00
59 T	2-Hexanone	0.103	0.098	4.9	106	0.00
60 T	Dibromochloromethane	0.300	0.298	0.7	110	0.00
61 T	1,2-Dibromoethane	0.222	0.210	5.4	104	0.00
62 S	4-Bromofluorobenzene	0.445	0.420	5.6	97	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	105	0.00
64 T	Tetrachloroethene	0.357	0.358	-0.3	112	0.00
65 PM	Chlorobenzene	0.981	0.958	2.3	107	0.00
66 T	1,1,1,2-Tetrachloroethane	0.353	0.353	0.0	110	0.00
67 C	Ethyl Benzene	1.830	1.777	2.9#	106	0.00
68 T	m/p-Xylenes	0.717	0.707	1.4	106	0.00
69 T	o-Xylene	0.669	0.660	1.3	106	0.00
70 T	Styrene	1.141	1.126	1.3	106	0.00
71 P	Bromoform	0.199	0.195	2.0	108	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	103	0.00
73 T	Isopropylbenzene	3.467	3.382	2.5	107	0.00
74 T	N-amyl acetate	0.603	0.534	11.4	96	0.00
75 P	1,1,2,2-Tetrachloroethane	0.492	0.446	9.3	103	0.00
76 T	1,2,3-Trichloropropane	0.387	0.347	10.3	101	0.00
77 T	Bromobenzene	0.796	0.798	-0.3	107	0.00
78 T	n-propylbenzene	4.036	3.868	4.2	104	0.00
79 T	2-Chlorotoluene	2.275	2.220	2.4	105	0.00
80 T	1,3,5-Trimethylbenzene	2.986	2.903	2.8	105	0.00
81 T	trans-1,4-Dichloro-2-butene	0.177	0.158	10.7	97	0.00
82 T	4-Chlorotoluene	2.415	2.342	3.0	104	0.00
83 T	tert-Butylbenzene	2.610	2.568	1.6	107	0.00
84 T	1,2,4-Trimethylbenzene	2.945	2.892	1.8	105	0.00
85 T	sec-Butylbenzene	3.605	3.468	3.8	104	0.00
86 T	p-Isopropyltoluene	3.349	3.271	2.3	103	0.00
87 T	1,3-Dichlorobenzene	1.583	1.568	0.9	106	0.00
88 T	1,4-Dichlorobenzene	1.551	1.559	-0.5	108	0.00
89 T	n-Butylbenzene	3.061	2.904	5.1	101	0.00

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW042420\
Data File : VW015333.D
Acq On : 24 Apr 2020 10:37
Operator : SY/VA
Sample : VSTDCCC050
Misc : 5.00G/5ML/MSVOA W/SOIL
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_W
LabSampleId :
VSTDCCC050

Quant Time: Apr 25 06:54:37 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W042020S.M
Quant Title : SW846 8260
QLast Update : Tue Apr 21 06:39:58 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.614	0.592	3.6	105	0.00
91 T	1,2-Dichlorobenzene	1.375	1.389	-1.0	108	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.089	0.081	9.0	100	0.00
93 T	1,2,4-Trichlorobenzene	0.956	0.929	2.8	103	0.00
94 T	Hexachlorobutadiene	0.637	0.619	2.8	107	0.00
95 T	Naphthalene	1.538	1.498	2.6	102	0.00
96 T	1,2,3-Trichlorobenzene	0.778	0.785	-0.9	107	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6