


Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW042721\
Data File : VW018802.D
Acq On : 27 Apr 2021 14:46
Operator : SY/VA
Sample : M2177-03
Misc : 5.33G/10ML/MSVOA_W/SOIL
ALS Vial : 8 Sample Multiplier: 1

Quant Time: Apr 28 00:42:41 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM041421SMA.M
Quant Title : SFAM01.0
QLast Update : Wed Apr 28 00:41:12 2021
Response via : Initial Calibration

Instrument :
MSVOA_W
ClientSampleId :
BG580

Manual Integrations
APPROVED

MMDadoda
4/29/2021 10:03:50 AM





Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW042721\
Data File : VW018802.D
Acq On : 27 Apr 2021 14:46
Operator : SY/VA
Sample : M2177-03
Misc : 5.33G/10ML/MSVOA_W/SOIL
ALS Vial : 8 Sample Multiplier: 1

Quant Time: May 15 06:31:23 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM041421SMA.M
Quant Title : SFAM01.0
QLast Update : Thu Apr 29 00:13:39 2021
Response via : Initial Calibration

Instrument :
MSVOA_W
ClientSampleId :
BG580

Manual Integrations
APPROVED

MMDadoda
4/29/2021 10:03:50 AM



(7) Chloroethane-d5 (S)

3.056min (+ 0.165) 0.03 ug/L

response 266

Ion	Exp%	Act%
69.00	100.00	100.00
71.00	32.10	24.81
51.00	28.40	23.68
0.00	0.00	0.00




Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW042721\
Data File : VW018802.D
Acq On : 27 Apr 2021 14:46
Operator : SY/VA
Sample : M2177-03
Misc : 5.33G/10ML/MSVOA_W/SOIL
ALS Vial : 8 Sample Multiplier: 1

Quant Time: Apr 28 00:42:41 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM041421SMA.M
Quant Title : SFAM01.0
Qlast Update : Wed Apr 28 00:41:12 2021
Response via : Initial Calibration

Instrument :
MSVOA_W
ClientSampleId :
BG580

Manual Integrations
APPROVED

MMDadoda
4/29/2021 10:03:50 AM



TIC: VW018802.D\data.ms

(7) Chloroethane-d5 (S)

2.897min (+ 0.006) 12.64 ug/L m

response 97673

Ion	Exp%	Act%
69.00	100.00	100.00
71.00	32.10	0.07#
51.00	28.40	0.06#
0.00	0.00	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW042721\
 Data File : VW018802.D
 Acq On : 27 Apr 2021 14:46
 Operator : SY/VA
 Sample : M2177-03
 Misc : 5.33G/10ML/MSVOA_W/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Quant Time: Apr 28 00:42:41 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM041421SMA.M
 Quant Title : SFAM01.0
 QLast Update : Wed Apr 28 00:41:12 2021
 Response via : Initial Calibration

Instrument :
 MSVOA_W
 ClientSampleId :
 BG580

Manual Integrations
 APPROVED

MMDadoda
 4/29/2021 10:03:50 AM

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	573523	25.00	ug/L	0.00
28) Chlorobenzene-d5	11.634	117	494268	25.00	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	213848	25.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.361	65	81828	9.39	ug/L	0.00
Spiked Amount 25.000	Range 30 - 150		Recovery =	37.56%		
7) Chloroethane-d5	2.897	69	97673m	12.64	ug/L	0.00
Spiked Amount 25.000	Range 30 - 150		Recovery =	50.56%		
11) 1,1-Dichloroethene-d2	4.019	63	136474	9.02	ug/L	0.00
Spiked Amount 25.000	Range 45 - 110		Recovery =	36.08%#		
21) 2-Butanone-d5	7.080	46	90142	45.44	ug/L	0.00
Spiked Amount 50.000	Range 20 - 135		Recovery =	90.88%		
24) Chloroform-d	7.647	84	307831	19.56	ug/L	0.00
Spiked Amount 25.000	Range 40 - 150		Recovery =	78.24%		
26) 1,2-Dichloroethane-d4	8.305	65	173802	19.64	ug/L	0.00
Spiked Amount 25.000	Range 70 - 130		Recovery =	78.56%		
32) Benzene-d6	8.275	84	509243	17.56	ug/L	0.00
Spiked Amount 25.000	Range 20 - 135		Recovery =	70.24%		
36) 1,2-Dichloropropane-d6	9.274	67	185548	21.23	ug/L	0.00
Spiked Amount 25.000	Range 70 - 120		Recovery =	84.92%		
41) Toluene-d8	10.323	98	425413	16.04	ug/L	0.00
Spiked Amount 25.000	Range 30 - 130		Recovery =	64.16%		
43) trans-1,3-Dichloroprop...	10.579	79	67200	18.06	ug/L	0.00
Spiked Amount 25.000	Range 30 - 135		Recovery =	72.24%		
47) 2-Hexanone-d5	10.927	63	78652	51.76	ug/L	0.00
Spiked Amount 50.000	Range 20 - 135		Recovery =	103.52%		
56) 1,1,2,2-Tetrachloroeth...	12.694	84	164790	22.58	ug/L	0.00
Spiked Amount 25.000	Range 45 - 120		Recovery =	90.32%		
66) 1,2-Dichlorobenzene-d4	13.853	152	165781	22.07	ug/L	0.00
Spiked Amount 25.000	Range 75 - 120		Recovery =	88.28%		
Target Compounds						
13) Acetone	4.129	43	6471	4.49	ug/L	90
14) Carbon disulfide	4.385	76	19744	0.88	ug/L	99
16) Methylene chloride	4.921	84	33048	3.52	ug/L	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

SY
 4/30/21