

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW042821\
 Data File : VW018818.D
 Acq On : 28 Apr 2021 14:13
 Operator : SY/VA
 Sample : M2166-05
 Misc : 5.91G/10ML/MSVOA_W/SOIL
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 GB3G8

Manual Integrations
 APPROVED

MMDadoda
 4/30/2021 10:48:06 AM

Quant Time: Apr 29 00:15:07 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM041421SMA.M
 Quant Title : SFAM01.0
 QLast Update : Thu Apr 29 00:13:39 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	635854	25.00	ug/L	0.00
28) Chlorobenzene-d5	11.634	117	544258	25.00	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.560	152	242386	25.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.361	65	97960	10.14	ug/L	0.00
Spiked Amount 25.000	Range 30 - 150		Recovery =	40.56%		
7) Chloroethane-d5	2.891	69	95770	11.18	ug/L	0.00
Spiked Amount 25.000	Range 30 - 150		Recovery =	44.72%		
11) 1,1-Dichloroethene-d2	4.025	63	138726	8.27	ug/L	0.00
Spiked Amount 25.000	Range 45 - 110		Recovery =	33.08%#		
21) 2-Butanone-d5	7.080	46	134989	61.37	ug/L	0.00
Spiked Amount 50.000	Range 20 - 135		Recovery =	122.74%		
24) Chloroform-d	7.647	84	335449	19.22	ug/L	0.00
Spiked Amount 25.000	Range 40 - 150		Recovery =	76.88%		
26) 1,2-Dichloroethane-d4	8.305	65	211057	21.51	ug/L	0.00
Spiked Amount 25.000	Range 70 - 130		Recovery =	86.04%		
32) Benzene-d6	8.275	84	525977	16.47	ug/L	0.00
Spiked Amount 25.000	Range 20 - 135		Recovery =	65.88%		
36) 1,2-Dichloropropane-d6	9.274	67	204001	21.20	ug/L	0.00
Spiked Amount 25.000	Range 70 - 120		Recovery =	84.80%		
41) Toluene-d8	10.323	98	427891	14.65	ug/L	0.00
Spiked Amount 25.000	Range 30 - 130		Recovery =	58.60%		
43) trans-1,3-Dichloroprop...	10.579	79	84933	20.73	ug/L	0.00
Spiked Amount 25.000	Range 30 - 135		Recovery =	82.92%		
47) 2-Hexanone-d5	10.927	63	119327	71.32	ug/L	0.00
Spiked Amount 50.000	Range 20 - 135		Recovery =	142.64%#		
56) 1,1,2,2-Tetrachloroeth...	12.695	84	232244	28.90	ug/L	0.00
Spiked Amount 25.000	Range 45 - 120		Recovery =	115.60%		
66) 1,2-Dichlorobenzene-d4	13.853	152	180824	21.24	ug/L	0.00
Spiked Amount 25.000	Range 75 - 120		Recovery =	84.96%		
Target Compounds						Qvalue
13) Acetone	4.135	43	39775	24.91	ug/L	90
15) Methyl Acetate	4.678	43	3681m	1.01	ug/L	
16) Methylene chloride	4.915	84	64476	6.19	ug/L	98
22) 2-Butanone	7.177	43	7068	2.91	ug/L	77
46) Tetrachloroethene	10.866	164	1123730	180.73	ug/L	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW042821\
Data File : VW018818.D
Acq On : 28 Apr 2021 14:13
Operator : SY/VA
Sample : M2166-05
Misc : 5.91G/10ML/MSVOA_W/SOIL
ALS Vial : 6 Sample Multiplier: 1


Instrument :
MSVOA_W
ClientSampleId :
GB3G8

Manual Integrations
APPROVED

MMDadoda
4/30/2021 10:48:06 AM

Quant Time: Apr 29 00:15:07 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM041421SMA.M
Quant Title : SFAM01.0
QLast Update : Thu Apr 29 00:13:39 2021
Response via : Initial Calibration









#13
Acetone
Concen: 24.91 ug/L
RT: 4.135 min Scan# 366
Delta R.T. 0.000 min
Lab File: VW018818.D
Acq: 28 Apr 2021 14:13

Tgt Ion: 43 Resp: 39775
Ion Ratio Lower Upper
43 100
58 32.2 0.0 53.8

Instrument :
MSVOA_W
ClientSampled :
GB3G8




Manual Integrations
APPROVED


MMDadoda
4/30/2021 10:48:06 AM



#15
Methyl Acetate
Concen: 1.01 ug/L m
RT: 4.678 min Scan# 455
Delta R.T. 0.006 min
Lab File: VW018818.D
Acq: 28 Apr 2021 14:13


Tgt Ion: 43 Resp: 3681
Ion Ratio Lower Upper
43 100
74 13.5 20.0 30.0#





#16
Methylene chloride
Concen: 6.19 ug/L
RT: 4.915 min Scan# 494
Delta R.T. 0.000 min
Lab File: VW018818.D
Acq: 28 Apr 2021 14:13

Instrument :
MSVOA_W
ClientSampleId :
GB3G8





Tgt Ion: 84 Resp: 64476


Ion	Ratio	Lower	Upper
84	100		
86	63.2	45.3	84.1
49	121.0	86.6	160.8

Manual Integrations
APPROVED

MMDadoda
4/30/2021 10:48:06 AM






#22
2-Butanone
Concen: 2.91 ug/L
RT: 7.177 min Scan# 865
Delta R.T. 0.006 min
Lab File: VW018818.D
Acq: 28 Apr 2021 14:13




Tgt Ion: 43 Resp: 7068

Ion	Ratio	Lower	Upper
43	100		
72	17.2	14.9	44.7







#46
Tetrachloroethene
Concen: 180.73 ug/L
RT: 10.866 min Scan# 1470
Delta R.T. 0.000 min
Lab File: VW018818.D
Acq: 28 Apr 2021 14:13



Tgt Ion:164 Resp: 1123730
Ion Ratio Lower Upper
164 100
129 97.5 66.7 123.9
131 85.5 61.1 113.5
166 114.1 87.0 161.6



Instrument :
MSVOA_W
ClientSampleId :
GB3G8

Manual Integrations
APPROVED

MMDadoda
4/30/2021 10:48:06 AM