

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW042822\
 Data File : VW022667.D
 Acq On : 28 Apr 2022 15:56
 Operator : SY/VA
 Sample : N2603-04
 Misc : 5.42g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 EW485

Quant Time: Apr 29 01:11:27 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM040722SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Apr 29 01:08:57 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.841	114	286832	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.634	117	238857	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	79831	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.355	65	72819	11.484	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	45.920%
7) Chloroethane-d5	2.891	69	59775	15.041	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	60.160%
11) 1,1-Dichloroethene-d2	4.019	63	82200	9.995	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	39.960%#
21) 2-Butanone-d5	7.080	46	64529	58.289	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	116.580%
24) Chloroform-d	7.653	84	178378	19.896	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	79.600%
26) 1,2-Dichloroethane-d4	8.305	65	109689	21.413	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	85.640%
32) Benzene-d6	8.274	84	296562	20.322	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	81.280%
36) 1,2-Dichloropropane-d6	9.274	67	101424	24.016	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	96.080%
41) Toluene-d8	10.323	98	239353	16.747	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	67.000%
43) trans-1,3-Dichloroprop...	10.579	79	34771	17.446	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	69.800%
47) 2-Hexanone-d5	10.920	63	50496	64.267	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	128.540%
56) 1,1,2,2-Tetrachloroeth...	12.688	84	109702	30.550	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	122.200%#
66) 1,2-Dichlorobenzene-d4	13.853	152	58945	20.846	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	83.400%
Target Compounds					Qvalue	
42) Toluene	10.390	91	17991	0.968	ug/L	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW042822\
Data File : VW022667.D
Acq On : 28 Apr 2022 15:56
Operator : SY/VA
Sample : N2603-04
Misc : 5.42g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
EW485

Quant Time: Apr 29 01:11:27 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM040722SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Apr 29 01:08:57 2022
Response via : Initial Calibration

