

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW042822\
 Data File : VW022684.D
 Acq On : 29 Apr 2022 00:47
 Operator : SY/VA
 Sample : N2603-07RE
 Misc : 6.14g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 EW488RE

Quant Time: Apr 29 02:22:38 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM040722SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Apr 29 02:06:41 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.841	114	246380	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.627	117	189756	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	55550	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.355	65	60964	11.193	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	44.760%
7) Chloroethane-d5	2.885	69	49707	14.562	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	58.240%
11) 1,1-Dichloroethene-d2	4.019	63	67534	9.560	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	38.240%#
21) 2-Butanone-d5	7.079	46	38765	40.766	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	81.540%
24) Chloroform-d	7.646	84	155501	20.192	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	80.760%
26) 1,2-Dichloroethane-d4	8.305	65	86026	19.551	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	78.200%
32) Benzene-d6	8.274	84	242576	20.924	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	83.680%
36) 1,2-Dichloropropane-d6	9.274	67	82975	24.732	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	98.920%
41) Toluene-d8	10.323	98	188533	16.604	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	66.400%
43) trans-1,3-Dichloroprop...	10.579	79	19222	12.140	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	48.560%
47) 2-Hexanone-d5	10.920	63	26416	42.319	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	84.640%
56) 1,1,2,2-Tetrachloroeth...	12.688	84	75806	26.573	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	106.280%
66) 1,2-Dichlorobenzene-d4	13.853	152	39051	19.847	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	79.400%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW042822\
Data File : VW022684.D
Acq On : 29 Apr 2022 00:47
Operator : SY/VA
Sample : N2603-07RE
Misc : 6.14g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
EW488RE

Quant Time: Apr 29 02:22:38 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM040722SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Apr 29 02:06:41 2022
Response via : Initial Calibration

