

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW043021\
 Data File : VW018873.D
 Acq On : 30 Apr 2021 11:54
 Operator : SY/VA
 Sample : M2177-20
 Misc : 5.72G/10ML/MSVOA_W/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 BG590

Quant Time: Apr 30 22:54:37 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM042921SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Apr 30 22:53:19 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Difluorobenzene	8.841	114	732472	25.00	ug/L	0.00
28) Chlorobenzene-d5	11.634	117	515124	25.00	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.560	152	166311	25.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.355	65	202231	21.52	ug/L	0.00
Spiked Amount 25.000	Range 30 - 150		Recovery =	86.08%		
7) Chloroethane-d5	2.891	69	165490	22.73	ug/L	0.00
Spiked Amount 25.000	Range 30 - 150		Recovery =	90.92%		
11) 1,1-Dichloroethene-d2	4.019	63	307899	16.50	ug/L	0.00
Spiked Amount 25.000	Range 45 - 110		Recovery =	66.00%		
21) 2-Butanone-d5	7.086	46	92580	37.83	ug/L	0.00
Spiked Amount 50.000	Range 20 - 135		Recovery =	75.66%		
24) Chloroform-d	7.653	84	406368	21.14	ug/L	0.00
Spiked Amount 25.000	Range 40 - 150		Recovery =	84.56%		
26) 1,2-Dichloroethane-d4	8.305	65	226217	21.07	ug/L	0.00
Spiked Amount 25.000	Range 70 - 130		Recovery =	84.28%		
32) Benzene-d6	8.275	84	816548	28.05	ug/L	0.00
Spiked Amount 25.000	Range 20 - 135		Recovery =	112.20%		
36) 1,2-Dichloropropane-d6	9.274	67	239028	27.68	ug/L	0.00
Spiked Amount 25.000	Range 70 - 120		Recovery =	110.72%		
41) Toluene-d8	10.323	98	699655	25.78	ug/L	0.00
Spiked Amount 25.000	Range 30 - 130		Recovery =	103.12%		
43) trans-1,3-Dichloroprop...	10.579	79	80733	20.57	ug/L	0.00
Spiked Amount 25.000	Range 30 - 135		Recovery =	82.28%		
47) 2-Hexanone-d5	10.926	63	80878	52.31	ug/L	0.00
Spiked Amount 50.000	Range 20 - 135		Recovery =	104.62%		
56) 1,1,2,2-Tetrachloroeth...	12.694	84	159106	21.75	ug/L	0.00
Spiked Amount 25.000	Range 45 - 120		Recovery =	87.00%		
66) 1,2-Dichlorobenzene-d4	13.853	152	121994	21.41	ug/L	0.00
Spiked Amount 25.000	Range 75 - 120		Recovery =	85.64%		
Target Compounds					Qvalue	
13) Acetone	4.135	43	21811	11.63	ug/L	89
14) Carbon disulfide	4.385	76	15596	0.56	ug/L	98
16) Methylene chloride	4.921	84	36370	3.00	ug/L	93
42) Toluene	10.390	91	15005	0.44	ug/L	95


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW043021\
Data File : VW018873.D
Acq On : 30 Apr 2021 11:54
Operator : SY/VA
Sample : M2177-20
Misc : 5.72G/10ML/MSVOA_W/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BG590

Quant Time: Apr 30 22:54:37 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM042921SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Apr 30 22:53:19 2021
Response via : Initial Calibration









#13
Acetone
Concen: 11.63 ug/L
RT: 4.135 min Scan# 366
Delta R.T. -0.006 min
Lab File: VW018873.D
Acq: 30 Apr 2021 11:54




Instrument :
MSVOA_W
ClientSampled :
BG590


Tgt Ion: 43 Resp: 21811
Ion Ratio Lower Upper
43 100
58 32.5 0.0 53.8



#14
Carbon disulfide
Concen: 0.56 ug/L
RT: 4.385 min Scan# 407
Delta R.T. -0.000 min
Lab File: VW018873.D
Acq: 30 Apr 2021 11:54

Tgt Ion: 76 Resp: 15596
Ion Ratio Lower Upper
76 100
78 9.8 7.4 11.0









#16
Methylene chloride
Concen: 3.00 ug/L
RT: 4.921 min Scan# 495
Delta R.T. 0.006 min
Lab File: VW018873.D
Acq: 30 Apr 2021 11:54

Instrument :
MSVOA_W
ClientSampled :
BG590

Tgt Ion: 84 Resp: 36370
Ion Ratio Lower Upper
84 100
86 66.6 45.3 84.1
49 112.4 86.6 160.8



#42
Toluene
Concen: 0.44 ug/L
RT: 10.390 min Scan# 1392
Delta R.T. -0.000 min
Lab File: VW018873.D
Acq: 30 Apr 2021 11:54

Tgt Ion: 91 Resp: 15005
Ion Ratio Lower Upper
91 100
92 52.6 39.5 73.3

