

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW043021\
 Data File : VW018886.D
 Acq On : 30 Apr 2021 17:09
 Operator : SY/VA
 Sample : M2209-03
 Misc : 5.38G/10ML/MSVOA_W/SOIL
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 BG7E2

Quant Time: Apr 30 22:57:04 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM042921SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Apr 30 22:53:19 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Difluorobenzene	8.847	114	665800	25.00	ug/L	0.00
28) Chlorobenzene-d5	11.633	117	601488	25.00	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.560	152	282919	25.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.361	65	187580	21.96	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	87.84%
7) Chloroethane-d5	2.897	69	149258	22.55	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	90.20%
11) 1,1-Dichloroethene-d2	4.019	63	288952	17.04	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	68.16%
21) 2-Butanone-d5	7.085	46	79612	35.79	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	71.58%
24) Chloroform-d	7.652	84	401175	22.96	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	91.84%
26) 1,2-Dichloroethane-d4	8.305	65	213230	21.85	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	87.40%
32) Benzene-d6	8.274	84	808620	23.79	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	95.16%
36) 1,2-Dichloropropane-d6	9.274	67	235461	23.35	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	93.40%
41) Toluene-d8	10.323	98	747047	23.57	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	94.28%
43) trans-1,3-Dichloroprop...	10.579	79	95413	20.82	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	83.28%
47) 2-Hexanone-d5	10.926	63	70019	38.79	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	77.58%
56) 1,1,2,2-Tetrachloroeth...	12.694	84	169645	19.86	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	79.44%
66) 1,2-Dichlorobenzene-d4	13.852	152	237875	24.54	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	98.16%
Target Compounds						
					Qvalue	
13) Acetone	4.135	43	6526	3.83	ug/L	92
16) Methylene chloride	4.921	84	53064	4.82	ug/L	93


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW043021\
Data File : VW018886.D
Acq On : 30 Apr 2021 17:09
Operator : SY/VA
Sample : M2209-03
Misc : 5.38G/10ML/MSVOA_W/SOIL
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BG7E2

Quant Time: Apr 30 22:57:04 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM042921SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Apr 30 22:53:19 2021
Response via : Initial Calibration









#13
Acetone
Concen: 3.83 ug/L
RT: 4.135 min Scan# 366
Delta R.T. -0.007 min
Lab File: VW018886.D
Acq: 30 Apr 2021 17:09

Instrument :
MSVOA_W
ClientSampled :
BG7E2

Tgt Ion: 43 Resp: 6526
Ion Ratio Lower Upper
43 100
58 31.1 0.0 53.8



#16
Methylene chloride
Concen: 4.82 ug/L
RT: 4.921 min Scan# 495
Delta R.T. 0.006 min
Lab File: VW018886.D
Acq: 30 Apr 2021 17:09

Tgt Ion: 84 Resp: 53064
Ion Ratio Lower Upper
84 100
86 66.7 45.3 84.1
49 113.4 86.6 160.8

