

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW050222\
 Data File : VW022744.D
 Acq On : 02 May 2022 17:04
 Operator : SY/VA
 Sample : N2639-02
 Misc : 6.64g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 EXKZ8

Quant Time: May 03 01:44:18 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM050222SMA.M
 Quant Title : SFAM01.0
 QLast Update : Tue May 03 01:41:22 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	156171	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.628	117	150154	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.560	152	71419	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.349	65	148912	46.476	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	185.920%#
7) Chloroethane-d5	2.885	69	93498	44.854	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	179.400%#
11) 1,1-Dichloroethene-d2	4.019	63	151108	36.499	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	146.000%#
21) 2-Butanone-d5	7.074	46	63542	104.304	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	208.600%#
24) Chloroform-d	7.647	84	215609	47.123	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	188.480%#
26) 1,2-Dichloroethane-d4	8.305	65	125094	49.849	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	199.400%#
32) Benzene-d6	8.275	84	430367	49.726	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	198.920%#
36) 1,2-Dichloropropane-d6	9.275	67	122101	50.224	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	200.880%#
41) Toluene-d8	10.323	98	408610	48.874	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	195.480%#
43) trans-1,3-Dichloroprop...	10.579	79	52336	46.754	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	187.000%#
47) 2-Hexanone-d5	10.921	63	47734	102.852	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	205.700%#
56) 1,1,2,2-Tetrachloroeth...	12.689	84	111297	49.116	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	196.480%#
66) 1,2-Dichlorobenzene-d4	13.853	152	120800	48.386	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	193.560%#
Target Compounds						Qvalue
46) Tetrachloroethene	10.860	164	22406	12.256	ug/L	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW050222\
Data File : VW022744.D
Acq On : 02 May 2022 17:04
Operator : SY/VA
Sample : N2639-02
Misc : 6.64g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
EXKZ8

Quant Time: May 03 01:44:18 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM050222SMA.M
Quant Title : SFAM01.0
QLast Update : Tue May 03 01:41:22 2022
Response via : Initial Calibration

