




Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW050222\
 Data File : VW022754.D
 Acq On : 02 May 2022 21:05
 Operator : SY/VA
 Sample : N2638-12
 Misc : 5.87g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 EXL05

Manual IntegrationsAPPROVED

Quant Time: May 03 01:46:08 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM050222SMA.M
 Quant Title : SFAM01.0
 QLast Update : Tue May 03 01:41:22 2022
 Response via : Initial Calibration

Reviewed By :Semsettin Yesilyurt 05/03/2022
 Supervised By :Mahesh Dadoda 05/09/2022



TIC: VW022754.D\data.ms

(13) Acetone (T)

4.129min (+ 0.006) 6.66 ug/L m

response 4797

Ion	Exp%	Act%
43.00	100.00	100.00
58.00	32.00	23.72
0.00	0.00	0.00
0.00	0.00	0.00




Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW050222\
 Data File : VW022754.D
 Acq On : 02 May 2022 21:05
 Operator : SY/VA
 Sample : N2638-12
 Misc : 5.87g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 EXL05

Manual IntegrationsAPPROVED

Quant Time: May 03 01:46:08 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM050222SMA.M
 Quant Title : SFAM01.0
 QLast Update : Tue May 03 01:41:22 2022
 Response via : Initial Calibration

Reviewed By :Semsettin Yesilyurt 05/03/2022
 Supervised By :Mahesh Dadoda 05/09/2022



TIC: VW022754.D\data.ms

(13) Acetone (T)

4.129min (+ 0.006) 5.18 ug/L

response 3732

Ion	Exp%	Act%
43.00	100.00	100.00
58.00	32.00	30.49
0.00	0.00	0.00
0.00	0.00	0.00




Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW050222\
Data File : VW022754.D
Acq On : 02 May 2022 21:05
Operator : SY/VA
Sample : N2638-12
Misc : 5.87g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
EXL05

Manual IntegrationsAPPROVED

Quant Time: May 03 01:46:08 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM050222SMA.M
Quant Title : SFAM01.0
QLast Update : Tue May 03 01:41:22 2022
Response via : Initial Calibration

Reviewed By :Semsettin Yesilyurt 05/03/2022
Supervised By :Mahesh Dadoda 05/09/2022



TIC: VW022754.D\data.ms

(26) 1,2-Dichloroethane-d4 (S)

8.457min (+ 0.153) 0.03 ug/L

response 107

Ion	Exp%	Act%
65.00	100.00	100.00
67.00	51.10	36.45
51.00	92.30	80.37
0.00	0.00	0.00




Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW050222\
 Data File : VW022754.D
 Acq On : 02 May 2022 21:05
 Operator : SY/VA
 Sample : N2638-12
 Misc : 5.87g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 EXL05

Manual IntegrationsAPPROVED

Reviewed By :Semsettin Yesilyurt 05/03/2022
 Supervised By :Mahesh Dadoda 05/09/2022

Quant Time: May 03 01:46:08 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM050222SMA.M
 Quant Title : SFAM01.0
 QLast Update : Tue May 03 01:41:22 2022
 Response via : Initial Calibration



TIC: VW022754.D\data.ms

(26) 1,2-Dichloroethane-d4 (S)

8.305min (+ 0.000) 22.19 ug/L m

response 88000

Ion	Exp%	Act%
65.00	100.00	100.00
67.00	51.10	0.04#
51.00	92.30	0.10#
0.00	0.00	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW050222\
 Data File : VW022754.D
 Acq On : 02 May 2022 21:05
 Operator : SY/VA
 Sample : N2638-12
 Misc : 5.87g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 EXL05

Manual IntegrationsAPPROVED

Reviewed By :Semsettin Yesilyurt 05/03/2022
 Supervised By :Mahesh Dadoda 05/09/2022

Quant Time: May 03 01:46:08 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM050222SMA.M
 Quant Title : SFAM01.0
 QLast Update : Tue May 03 01:41:22 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	246790	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.628	117	230944	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	104298	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.349	65	113545	22.425	ug/L	0.00
Spiked Amount 25.000	Range 30	- 150	Recovery	=	89.720%	
7) Chloroethane-d5	2.879	69	70287	21.338	ug/L	0.00
Spiked Amount 25.000	Range 30	- 150	Recovery	=	85.360%	
11) 1,1-Dichloroethene-d2	4.013	63	105657	16.150	ug/L	0.00
Spiked Amount 25.000	Range 45	- 110	Recovery	=	64.600%	
21) 2-Butanone-d5	7.074	46	42122	43.755	ug/L	0.00
Spiked Amount 50.000	Range 20	- 135	Recovery	=	87.500%	
24) Chloroform-d	7.653	84	154237	21.332	ug/L	0.00
Spiked Amount 25.000	Range 40	- 150	Recovery	=	85.320%	
26) 1,2-Dichloroethane-d4	8.305	65	88000m	22.191	ug/L	0.00
Spiked Amount 25.000	Range 70	- 130	Recovery	=	88.760%	
32) Benzene-d6	8.275	84	306642	23.036	ug/L	0.00
Spiked Amount 25.000	Range 20	- 135	Recovery	=	92.160%	
36) 1,2-Dichloropropane-d6	9.274	67	86861	23.230	ug/L	0.00
Spiked Amount 25.000	Range 70	- 120	Recovery	=	92.920%	
41) Toluene-d8	10.323	98	283777	22.068	ug/L	0.00
Spiked Amount 25.000	Range 30	- 130	Recovery	=	88.280%	
43) trans-1,3-Dichloroprop...	10.573	79	35171	20.428	ug/L	0.00
Spiked Amount 25.000	Range 30	- 135	Recovery	=	81.720%	
47) 2-Hexanone-d5	10.920	63	32438	45.443	ug/L	0.00
Spiked Amount 50.000	Range 20	- 135	Recovery	=	90.880%	
56) 1,1,2,2-Tetrachloroeth...	12.688	84	77526	22.244	ug/L	0.00
Spiked Amount 25.000	Range 45	- 120	Recovery	=	88.960%	
66) 1,2-Dichlorobenzene-d4	13.853	152	79672	21.852	ug/L	0.00
Spiked Amount 25.000	Range 75	- 120	Recovery	=	87.400%	
Target Compounds						
						Qvalue
13) Acetone	4.129	43	4797m	6.656	ug/L	
22) 2-Butanone	7.177	43	8802	8.732	ug/L	94
46) Tetrachloroethene	10.859	164	52274	18.591	ug/L	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW050222\
Data File : VW022754.D
Acq On : 02 May 2022 21:05
Operator : SY/VA
Sample : N2638-12
Misc : 5.87g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
EXL05

Manual IntegrationsAPPROVED

Reviewed By :Semsettin Yesilyurt 05/03/2022
Supervised By :Mahesh Dadoda 05/09/2022

Quant Time: May 03 01:46:08 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM050222SMA.M
Quant Title : SFAM01.0
QLast Update : Tue May 03 01:41:22 2022
Response via : Initial Calibration

